

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE062519\
 Data File : BE100181.D
 Acq On : 25 Jun 2019 15:09
 Operator : JU/SJ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE062519

Quant Time: Jun 25 15:58:24 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE062519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 25 15:40:27 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	74	0.00
2	1,4-Dioxane	0.626	0.632	-1.0	69	0.00
3	n-Nitrosodimethylamine	0.232	0.203	12.5	78	0.01
4 S	2-Fluorophenol	0.889	0.794	10.7	64	0.00
5 S	Phenol-d6	1.223	1.081	11.6	69	0.02
6	bis(2-Chloroethyl)ether	1.120	1.008	10.0	73	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	79	0.00
8 S	Nitrobenzene-d5	0.398	0.343	13.8	68	0.01
9	Naphthalene	4.365	4.345	0.5	75	0.01
10	Hexachlorobutadiene	0.451	0.454	-0.7	74	0.01
11 SURR	2-Methylnaphthalene-d10	0.750	0.737	1.7	74	0.01
12	2-Methylnaphthalene	0.703	0.665	5.4	70	0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	76	0.01
14 S	2,4,6-Tribromophenol	0.232	0.197	15.1	67	0.01
15 S	2-Fluorobiphenyl	1.604	1.638	-2.1	75	0.01
16	Acenaphthylene	1.950	1.966	-0.8	72	0.01
17	Acenaphthene	1.093	1.111	-1.6	72	0.00
18	Fluorene	1.581	1.568	0.8	70	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	73	0.01
20	4-Bromophenyl-phenylether	0.248	0.248	0.0	70	0.01
21	Hexachlorobenzene	0.272	0.279	-2.6	70	0.00
22	Pentachlorophenol	0.064	0.043	32.8#	63	0.04
23	Phenanthrene	0.947	0.973	-2.7	70	0.01
24	Anthracene	0.800	0.761	4.9	69	0.01
25 SURR	Fluoranthene-d10	5.765	6.027	-4.5	74	0.00
26	Fluoranthene	1.541	1.624	-5.4	75	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	86	0.00
28	Pyrene	1.024	0.977	4.6	73	0.00
29 S	Terphenyl-d14	0.785	0.737	6.1	73	0.00
30	Benzo(a)anthracene	1.179	1.132	4.0	78	0.00
31	Chrysene	1.383	1.419	-2.6	85	0.00
32	Bis(2-ethylhexyl)phthalate	1.355	1.173	13.4	79	0.00
33	Indeno(1,2,3-cd)pyrene	1.969	1.925	2.2	91	0.00
34 I	Perylene-d12	1.000	1.000	0.0	96	0.00
35	Benzo(b)fluoranthene	1.076	1.056	1.9	89	0.00
36	Benzo(k)fluoranthene	1.287	1.305	-1.4	89	0.00
37 C	Benzo(a)pyrene	1.115	1.104	1.0	90	0.00
38	Dibenzo(a,h)anthracene	1.192	1.110	6.9	92	0.00
39	Benzo(g,h,i)perylene	1.245	1.209	2.9	92	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE062519\
Data File : BE100181.D
Acq On : 25 Jun 2019 15:09
Operator : JU/SJ
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE062519

Quant Time: Jun 25 15:58:24 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE062519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 25 15:40:27 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			