

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE062519\
 Data File : BE100181.D
 Acq On : 25 Jun 2019 15:09
 Operator : JU/SJ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE062519

Quant Time: Jun 25 15:58:24 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE062519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 25 15:40:27 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	74	0.00
2	1,4-Dioxane	0.400	0.404	-1.0	69	0.00
3	n-Nitrosodimethylamine	0.400	0.351	12.3	78	0.01
4 S	2-Fluorophenol	0.400	0.357	10.8	64	0.00
5 S	Phenol-d6	0.400	0.353	11.8	69	0.02
6	bis(2-Chloroethyl)ether	0.400	0.360	10.0	73	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	79	0.00
8 S	Nitrobenzene-d5	0.400	0.345	13.8	68	0.01
9	Naphthalene	0.400	0.398	0.5	75	0.01
10	Hexachlorobutadiene	0.400	0.403	-0.8	74	0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.393	1.8	74	0.01
12	2-Methylnaphthalene	0.400	0.379	5.3	70	0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	76	0.01
14 S	2,4,6-Tribromophenol	0.400	0.341	14.8	67	0.01
15 S	2-Fluorobiphenyl	0.400	0.409	-2.2	75	0.01
16	Acenaphthylene	0.400	0.403	-0.8	72	0.01
17	Acenaphthene	0.400	0.407	-1.7	72	0.00
18	Fluorene	0.400	0.397	0.8	70	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	73	0.01
20	4-Bromophenyl-phenylether	0.400	0.400	0.0	70	0.01
21	Hexachlorobenzene	0.400	0.410	-2.5	70	0.00
22	Pentachlorophenol	0.400	0.461	-15.3	63	0.04
23	Phenanthrene	0.400	0.411	-2.7	70	0.01
24	Anthracene	0.400	0.380	5.0	69	0.01
25 SURR	Fluoranthene-d10	0.400	0.418	-4.5	74	0.00
26	Fluoranthene	0.400	0.421	-5.2	75	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	86	0.00
28	Pyrene	0.400	0.382	4.5	73	0.00
29 S	Terphenyl-d14	0.400	0.375	6.3	73	0.00
30	Benzo(a)anthracene	0.400	0.384	4.0	78	0.00
31	Chrysene	0.400	0.410	-2.5	85	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.346	13.5	79	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.391	2.3	91	0.00
34 I	Perylene-d12	0.400	0.400	0.0	96	0.00
35	Benzo(b)fluoranthene	0.400	0.393	1.8	89	0.00
36	Benzo(k)fluoranthene	0.400	0.405	-1.3	89	0.00
37 C	Benzo(a)pyrene	0.400	0.396	1.0	90	0.00
38	Dibenzo(a,h)anthracene	0.400	0.372	7.0	92	0.00
39	Benzo(g,h,i)perylene	0.400	0.388	3.0	92	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE062519\
Data File : BE100181.D
Acq On : 25 Jun 2019 15:09
Operator : JU/SJ
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE062519

Quant Time: Jun 25 15:58:24 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE062519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 25 15:40:27 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				