

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE062519\
 Data File : BE100185.D
 Acq On : 25 Jun 2019 18:32
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 26 03:21:16 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE062519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 25 15:40:27 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2	1,4-Dioxane	0.626	0.657	-5.0	99	0.00
3	n-Nitrosodimethylamine	0.232	0.191	17.7	101	-0.03
4 S	2-Fluorophenol	0.889	0.900	-1.2	100	0.00
5 S	Phenol-d6	1.223	1.237	-1.1	110	0.00
6	bis(2-Chloroethyl)ether	1.120	0.829	26.0#	83	-0.03
7 I	Naphthalene-d8	1.000	1.000	0.0	105	0.00
8 S	Nitrobenzene-d5	0.398	0.367	7.8	97	0.00
9	Naphthalene	4.365	4.448	-1.9	102	0.01
10	Hexachlorobutadiene	0.451	0.459	-1.8	100	0.00
11 SURR	2-Methylnaphthalene-d10	0.750	0.740	1.3	99	0.00
12	2-Methylnaphthalene	0.703	0.694	1.3	97	0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	97	0.01
14 S	2,4,6-Tribromophenol	0.232	0.239	-3.0	104	0.00
15 S	2-Fluorobiphenyl	1.604	1.714	-6.9	100	0.01
16	Acenaphthylene	1.950	2.023	-3.7	95	0.00
17	Acenaphthene	1.093	1.158	-5.9	95	0.00
18	Fluorene	1.581	1.690	-6.9	96	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	92	0.00
20	4-Bromophenyl-phenylether	0.248	0.265	-6.9	93	0.00
21	Hexachlorobenzene	0.272	0.297	-9.2	93	0.00
22	Pentachlorophenol	0.064	0.072	-12.5	129	0.00
23	Phenanthrene	0.947	1.028	-8.6	93	0.00
24	Anthracene	0.800	0.830	-3.7	93	0.00
25 SURR	Fluoranthene-d10	5.765	6.101	-5.8	93	0.00
26	Fluoranthene	1.541	1.656	-7.5	95	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	102	0.00
28	Pyrene	1.024	1.081	-5.6	97	0.00
29 S	Terphenyl-d14	0.785	0.807	-2.8	95	0.00
30	Benzo(a)anthracene	1.179	1.260	-6.9	103	0.00
31	Chrysene	1.383	1.466	-6.0	105	0.00
32	Bis(2-ethylhexyl)phthalate	1.355	1.259	7.1	102	0.00
33	Indeno(1,2,3-cd)pyrene	1.969	2.287	-16.2	129	0.00
34 I	Perylene-d12	1.000	1.000	0.0	120	0.00
35	Benzo(b)fluoranthene	1.076	1.165	-8.3	123	0.00
36	Benzo(k)fluoranthene	1.287	1.295	-0.6	111	0.00
37 C	Benzo(a)pyrene	1.115	1.210	-8.5	124	0.00
38	Dibenzo(a,h)anthracene	1.192	1.248	-4.7	130	0.00
39	Benzo(g,h,i)perylene	1.245	1.296	-4.1	124	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE062519\
Data File : BE100185.D
Acq On : 25 Jun 2019 18:32
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 26 03:21:16 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE062519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 25 15:40:27 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			