

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE062618\
 Data File : BE096848.D
 Acq On : 26 Jun 2018 12:20
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 27 02:26:13 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE062018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 20 15:45:14 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	109	0.00
2	1,4-Dioxane	0.599	0.551	8.0	96	0.00
3	n-Nitrosodimethylamine	0.635	0.605	4.7	108	0.00
4 S	2-Fluorophenol	0.894	0.783	12.4	103	0.00
5 S	Phenol-d6	1.334	1.189	10.9	105	0.00
6	bis(2-Chloroethyl)ether	1.462	1.400	4.2	108	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	112	0.00
8 S	Nitrobenzene-d5	0.242	0.218	9.9	108	0.00
9	Naphthalene	3.630	3.559	2.0	110	0.00
10	Hexachlorobutadiene	0.151	0.155	-2.6	115	0.00
11 SURR	2-Methylnaphthalene-d10	0.576	0.555	3.6	108	0.00
12	2-Methylnaphthalene	0.602	0.587	2.5	110	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	108	0.00
14 S	2,4,6-Tribromophenol	0.095	0.074	22.1	98	0.00
15 S	2-Fluorobiphenyl	1.464	1.483	-1.3	108	0.00
16	Acenaphthylene	1.734	1.667	3.9	107	0.00
17	Acenaphthene	1.147	1.122	2.2	105	0.00
18	Fluorene	1.348	1.289	4.4	106	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	106	0.00
20	4-Bromophenyl-phenylether	0.180	0.178	1.1	106	0.00
21	Hexachlorobenzene	0.206	0.210	-1.9	105	-0.01
22	Pentachlorophenol	0.038	0.026	31.6#	95	0.00
23	Phenanthrene	0.983	0.981	0.2	106	0.00
24	Anthracene	0.815	0.773	5.2	106	0.00
25 SURR	Fluoranthene-d10	3.521	3.285	6.7	103	0.00
26	Fluoranthene	0.994	0.992	0.2	104	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	105	0.00
28	Pyrene	1.157	1.148	0.8	106	0.00
29 S	Terphenyl-d14	0.772	0.760	1.6	105	0.00
30	Benzo(a)anthracene	0.886	0.804	9.3	104	0.00
31	Chrysene	1.095	1.036	5.4	100	0.00
32	Bis(2-ethylhexyl)phthalate	0.794	0.662	16.6	97	0.00
33	Indeno(1,2,3-cd)pyrene	0.814	0.757	7.0	105	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	121	0.00
35	Benzo(b)fluoranthene	1.092	0.849	22.3	99	0.00
36	Benzo(k)fluoranthene	1.233	1.027	16.7	108	0.00
37 C	Benzo(a)pyrene	0.867	0.709	18.2	105	0.00
38	Dibenzo(a,h)anthracene	0.889	0.685	22.9	102	-0.01
39	Benzo(g,h,i)perylene	0.969	0.794	18.1	107	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE062618\
Data File : BE096848.D
Acq On : 26 Jun 2018 12:20
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 27 02:26:13 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE062018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 20 15:45:14 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			