

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE062619\
 Data File : BE100187.D
 Acq On : 26 Jun 2019 8:42
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 27 09:54:07 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE062519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 25 15:40:27 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	105	0.00
2	1,4-Dioxane	0.400	0.395	1.3	96	0.01
3	n-Nitrosodimethylamine	0.400	0.318	20.5	100	-0.01
4 S	2-Fluorophenol	0.400	0.415	-3.7	105	0.00
5 S	Phenol-d6	0.400	0.357	10.8	99	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.361	9.8	104	-0.02
7 I	Naphthalene-d8	0.400	0.400	0.0	105	0.00
8 S	Nitrobenzene-d5	0.400	0.381	4.8	100	0.00
9	Naphthalene	0.400	0.421	-5.2	105	0.00
10	Hexachlorobutadiene	0.400	0.448	-12.0	109	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.412	-3.0	103	0.00
12	2-Methylnaphthalene	0.400	0.423	-5.7	103	0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	102	0.00
14 S	2,4,6-Tribromophenol	0.400	0.380	5.0	101	0.00
15 S	2-Fluorobiphenyl	0.400	0.434	-8.5	107	0.01
16	Acenaphthylene	0.400	0.429	-7.2	103	0.00
17	Acenaphthene	0.400	0.425	-6.2	101	0.00
18	Fluorene	0.400	0.444	-11.0	105	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	104	0.00
20	4-Bromophenyl-phenylether	0.400	0.421	-5.2	104	0.00
21	Hexachlorobenzene	0.400	0.447	-11.7	108	0.00
22	Pentachlorophenol	0.400	0.471	-17.7	93	0.00
23	Phenanthrene	0.400	0.422	-5.5	102	0.00
24	Anthracene	0.400	0.397	0.8	101	0.00
25 SURR	Fluoranthene-d10	0.400	0.405	-1.3	101	0.00
26	Fluoranthene	0.400	0.412	-3.0	103	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	100	0.00
28	Pyrene	0.400	0.449	-12.2	101	0.00
29 S	Terphenyl-d14	0.400	0.439	-9.7	100	0.00
30	Benzo(a)anthracene	0.400	0.417	-4.2	99	-0.01
31	Chrysene	0.400	0.415	-3.7	100	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.332	17.0	89	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.345	13.8	94	0.00
34 I	Perylene-d12	0.400	0.400	0.0	92	0.00
35	Benzo(b)fluoranthene	0.400	0.438	-9.5	95	0.00
36	Benzo(k)fluoranthene	0.400	0.452	-13.0	95	0.00
37 C	Benzo(a)pyrene	0.400	0.418	-4.5	91	0.00
38	Dibenzo(a,h)anthracene	0.400	0.400	0.0	95	0.00
39	Benzo(g,h,i)perylene	0.400	0.394	1.5	90	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE062619\
Data File : BE100187.D
Acq On : 26 Jun 2019 8:42
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 27 09:54:07 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE062519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 25 15:40:27 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				