

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE062619\
 Data File : BE100196.D
 Acq On : 26 Jun 2019 17:31
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Jun 27 05:27:07 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE062519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 25 15:40:27 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	162#	0.00
2	1,4-Dioxane	0.626	0.528	15.7	126	0.00
3	n-Nitrosodimethylamine	0.232	0.298	-28.4#	248#	-0.05
4 S	2-Fluorophenol	0.889	1.040	-17.0	181#	0.00
5 S	Phenol-d6	1.223	1.157	5.4	162#	-0.02
6	bis(2-Chloroethyl)ether	1.120	1.102	1.6	173#	-0.04
7 I	Naphthalene-d8	1.000	1.000	0.0	161#	-0.01
8 S	Nitrobenzene-d5	0.398	0.310	22.1	126	-0.01
9	Naphthalene	4.365	3.888	10.9	136	0.00
10	Hexachlorobutadiene	0.451	0.369	18.2	123	0.00
11 SURR	2-Methylnaphthalene-d10	0.750	0.654	12.8	134	0.00
12	2-Methylnaphthalene	0.703	0.641	8.8	137	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	163#	0.00
14 S	2,4,6-Tribromophenol	0.232	0.175	24.6	128	0.00
15 S	2-Fluorobiphenyl	1.604	1.379	14.0	135	0.00
16	Acenaphthylene	1.950	1.713	12.2	135	0.00
17	Acenaphthene	1.093	0.978	10.5	136	0.00
18	Fluorene	1.581	1.446	8.5	138	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	164#	0.00
20	4-Bromophenyl-phenylether	0.248	0.212	14.5	134	0.00
21	Hexachlorobenzene	0.272	0.230	15.4	129	0.00
22	Pentachlorophenol	0.064	0.035	45.3#	112	0.00
23	Phenanthrene	0.947	0.850	10.2	137	0.00
24	Anthracene	0.800	0.751	6.1	151#	0.00
25 SURR	Fluoranthene-d10	5.765	5.136	10.9	140	0.00
26	Fluoranthene	1.541	1.350	12.4	138	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	156#	0.00
28	Pyrene	1.024	1.068	-4.3	145	0.00
29 S	Terphenyl-d14	0.785	0.805	-2.5	144	0.00
30	Benzo(a)anthracene	1.179	1.205	-2.2	150	-0.01
31	Chrysene	1.383	1.173	15.2	127	0.00
32	Bis(2-ethylhexyl)phthalate	1.355	1.623	-19.8	199#	0.00
33	Indeno(1,2,3-cd)pyrene	1.969	1.393	29.3#	120	0.00
34 I	Perylene-d12	1.000	1.000	0.0	151#	0.00
35	Benzo(b)fluoranthene	1.076	0.965	10.3	128	0.00
36	Benzo(k)fluoranthene	1.287	1.119	13.1	120	0.00
37 C	Benzo(a)pyrene	1.115	0.977	12.4	126	0.00
38	Dibenzo(a,h)anthracene	1.192	0.957	19.7	125	0.00
39	Benzo(g,h,i)perylene	1.245	1.006	19.2	121	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE062619\
Data File : BE100196.D
Acq On : 26 Jun 2019 17:31
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 27 05:27:07 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE062519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 25 15:40:27 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		