

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE062819\
 Data File : BE100251.D
 Acq On : 29 Jun 2019 1:05
 Operator : JU/SJ
 Sample : K3445-02
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 EXT-001-SW079-061719

Quant Time: Jun 29 03:35:51 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE062819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jun 29 00:22:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.65	152	780	0.40	ng	0.00
7) Naphthalene-d8	10.43	136	2765	0.40	ng	-0.01
13) Acenaphthene-d10	14.31	164	2162	0.40	ng	0.00
19) Phenanthrene-d10	17.05	188	6651	0.40	ng	0.00
27) Chrysene-d12	21.25	240	8956	0.40	ng	0.00
34) Perylene-d12	23.67	264	10772	0.40	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.22	112	624	0.31	ng	0.00
5) Phenol-d6	6.83	99	818	0.31	ng	-0.01
8) Nitrobenzene-d5	8.81	82	884	0.32	ng	-0.01
11) 2-Methylnaphthalene-d10	12.04	152	1935	0.37	ng	-0.01
14) 2,4,6-Tribromophenol	15.81	330	534	0.35	ng	-0.01
15) 2-Fluorobiphenyl	12.93	172	2683	0.32	ng	-0.01
25) Fluoranthene-d10	19.10	212	34447	0.36	ng	0.00
29) Terphenyl-d14	19.70	244	6132	0.33	ng	0.00


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) Naphthalene	10.48	128	1802	0.059	ng	# 91
12) 2-Methylnaphthalene	12.11	142	339	0.063	ng	# 88
16) Acenaphthylene	14.04	152	3362	0.324	ng	98
17) Acenaphthene	14.38	154	213	0.036	ng	# 86
18) Fluorene	15.35	166	648	0.075	ng	# 68
23) Phenanthrene	17.09	178	21367	1.325	ng	99
24) Anthracene	17.19	178	3649	0.251	ng	# 94
26) Fluoranthene	19.13	202	53836	2.097	ng	99
28) Pyrene	19.49	202	68428	2.886	ng	97
30) Benzo(a)anthracene	21.22	228	33094	1.124	ng	95
31) Chrysene	21.28	228	43447	1.469	ng	97
32) Bis(2-ethylhexyl)phthalate	21.15	149	25167	0.547	ng	98
33) Indeno(1,2,3-cd)pyrene	26.22	276	25528	0.653	ng	98
35) Benzo(b)fluoranthene	22.93	252	44293	1.517	ng	99
36) Benzo(k)fluoranthene	22.93	252	44293	1.359	ng	# 95
37) Benzo(a)pyrene	23.56	252	33427	1.142	ng	# 98
38) Dibenzo(a,h)anthracene	26.23	278	6655	0.218	ng	# 80
39) Benzo(g,h,i)perylene	27.02	276	28021	0.892	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE062819\
Data File : BE100251.D
Acq On : 29 Jun 2019 1:05
Operator : JU/SJ
Sample : K3445-02
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
EXT-001-SW079-061719

Quant Time: Jun 29 03:35:51 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE062819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jun 29 00:22:29 2019
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 0.400 ng

RT: 7.65 min Scan# 415

Delta R.T. 0.00 min

Lab File: BE100251.D

Acq: 29 Jun 2019 1:05

Instrument :

BNA_E

ClientSampleId :

EXT-001-SW079-061719


Tgt Ion:152 Resp: 780

Ion Ratio Lower Upper

152 100

150 132.0 114.5 171.7




115 65.0 57.1 85.7



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#4

2-Fluorophenol

Concen: 0.314 ng

RT: 5.22 min Scan# 204

Delta R.T. 0.00 min

Lab File: BE100251.D

Acq: 29 Jun 2019 1:05


Tgt Ion:112 Resp: 624

Ion Ratio Lower Upper

112 100

64 56.9 39.4 59.0



63 147.9 104.5 156.7



Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BE1

Ion 63.00 (62.70 to 63.70): BE1





#5

Phenol-d6

Concen: 0.313 ng

RT: 6.83 min Scan# 344

Delta R.T. -0.01 min

Lab File: BE100251.D

Acq: 29 Jun 2019 1:05

Instrument :

BNA_E

ClientSampleId :

EXT-001-SW079-061719


Tgt Ion: 99 Resp: 818

Ion Ratio Lower Upper




99 100

42 15.4 11.2 16.8

71 38.4 37.2 55.8



Abundance Ion 99.00 (98.70 to 99.70): BE1
Ion 42.00 (41.70 to 42.70): BE1
Ion 71.00 (70.70 to 71.70): BE1



#7

Naphthalene-d8

Concen: 0.400 ng

RT: 10.43 min Scan# 636

Delta R.T. -0.01 min

Lab File: BE100251.D

Acq: 29 Jun 2019 1:05

Tgt Ion: 136 Resp: 2765


Ion Ratio Lower Upper

136 100



137 14.1 11.8 17.8

54 33.8 27.7 41.5

68 9.0 6.8 10.2



Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BE1
Ion 68.00 (67.70 to 68.70): BE1





#8

Nitrobenzene-d5

Concen: 0.322 ng

RT: 8.81 min Scan# 511

Delta R.T. -0.01 min

Lab File: BE100251.D

Acq: 29 Jun 2019 1:05

Instrument :

BNA_E

ClientSampleId :

EXT-001-SW079-061719


Tgt Ion: 82 Resp: 884

Ion Ratio Lower Upper

82 100

128 157.4 140.9 211.3

54 173.7 146.2 219.2






Abundance Ion 82.00 (81.70 to 82.70): BE1

Ion 128.00 (127.70 to 128.70): BE1

Ion 54.00 (53.70 to 54.70): BE1

8.81

Time-->



#9

Naphthalene

Concen: 0.059 ng

RT: 10.48 min Scan# 640

Delta R.T. 0.00 min

Lab File: BE100251.D

Acq: 29 Jun 2019 1:05


Tgt Ion: 128 Resp: 1802

Ion Ratio Lower Upper

128 100

129 6.9 3.0 4.6#

127 7.5 3.5 5.3#





Abundance Ion 128.00 (127.70 to 128.70): E


Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E

10.48

Time-->









#11
2-Methylnaphthalene-d10
Concen: 0.372 ng
RT: 12.04 min Scan# 751
Delta R.T. -0.01 min
Lab File: BE100251.D
Acq: 29 Jun 2019 1:05




Instrument :
BNA_E
ClientSampleId :
EXT-001-SW079-061719

Tgt Ion:152 Resp: 1935
Ion Ratio Lower Upper
152 100
151 18.6 15.1 22.7



#12
2-Methylnaphthalene
Concen: 0.063 ng
RT: 12.11 min Scan# 756
Delta R.T. -0.01 min
Lab File: BE100251.D
Acq: 29 Jun 2019 1:05

Tgt Ion:142 Resp: 339
Ion Ratio Lower Upper
142 100
141 97.5 72.4 108.6
115 65.0 40.4 60.6#





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.31 min Scan# 909

Delta R.T. 0.00 min

Lab File: BE100251.D

Acq: 29 Jun 2019 1:05

Instrument :

BNA_E

ClientSampleId :

EXT-001-SW079-061719


Tgt Ion:164 Resp: 2162

Ion Ratio Lower Upper

164 100

162 101.6 83.4 125.0




160 50.5 45.0 67.6



Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#14

2,4,6-Tribromophenol

Concen: 0.353 ng

RT: 15.81 min Scan# 1014

Delta R.T. -0.01 min

Lab File: BE100251.D

Acq: 29 Jun 2019 1:05


Tgt Ion:330 Resp: 534

Ion Ratio Lower Upper

330 100

332 93.6 77.0 115.4



141 37.1 22.0 33.0#



Abundance Ion 330.00 (329.70 to 330.70): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E









#15
2-Fluorobiphenyl
Concen: 0.318 ng
RT: 12.93 min Scan# 813
Delta R.T. -0.01 min
Lab File: BE100251.D
Acq: 29 Jun 2019 1:05

Instrument :
BNA_E
ClientSampleId :
EXT-001-SW079-061719

Tgt Ion:172 Resp: 2683
Ion Ratio Lower Upper
172 100
171 38.1 31.9 47.9
170 28.9 22.2 33.2




Abundance Ion 172.00 (171.70 to 172.70): E
Ion 171.00 (170.70 to 171.70): E
Ion 170.00 (169.70 to 170.70): E






#16
Acenaphthylene
Concen: 0.324 ng
RT: 14.04 min Scan# 890
Delta R.T. 0.00 min
Lab File: BE100251.D
Acq: 29 Jun 2019 1:05

Tgt Ion:152 Resp: 3362
Ion Ratio Lower Upper
152 100
151 19.2 16.6 25.0
153 13.5 10.5 15.7



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 151.00 (150.70 to 151.70): E
Ion 153.00 (152.70 to 153.70): E





#17

Acenaphthene

Concen: 0.036 ng

RT: 14.38 min Scan# 914

Delta R.T. 0.00 min

Lab File: BE100251.D

Acq: 29 Jun 2019 1:05

Instrument :

BNA_E

ClientSampleId :

EXT-001-SW079-061719


Tgt Ion:154 Resp: 213

Ion Ratio Lower Upper

154 100

153 111.7 93.4 140.2




152 85.4 49.4 74.2#



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#18

Fluorene

Concen: 0.075 ng

RT: 15.35 min Scan# 980

Delta R.T. -0.00 min

Lab File: BE100251.D

Acq: 29 Jun 2019 1:05


Tgt Ion:166 Resp: 648

Ion Ratio Lower Upper

166 100

165 135.0 81.1 121.7#



167 23.0 10.7 16.1#




Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.05 min Scan# 1107

Delta R.T. -0.00 min

Lab File: BE100251.D

Acq: 29 Jun 2019 1:05

Instrument :

BNA_E

ClientSampleId :

EXT-001-SW079-061719


Tgt Ion:188 Resp: 6651

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0




80 8.4 7.6 11.4



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BE1

Ion 80.00 (79.70 to 80.70): BE1



#23

Phenanthrene

Concen: 1.325 ng

RT: 17.09 min Scan# 1110

Delta R.T. -0.01 min

Lab File: BE100251.D

Acq: 29 Jun 2019 1:05


Tgt Ion:178 Resp: 21367

Ion Ratio Lower Upper

178 100

176 19.5 15.9 23.9



179 15.9 12.2 18.2




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E





#24

Anthracene

Concen: 0.251 ng

RT: 17.19 min Scan# 1117

Delta R.T. -0.01 min

Lab File: BE100251.D

Acq: 29 Jun 2019 1:05

Instrument :

BNA_E

ClientSampleId :

EXT-001-SW079-061719


Tgt Ion:178 Resp: 3649

Ion Ratio Lower Upper

178 100

176 19.0 15.3 22.9




179 9.5 12.2 18.2#



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#25

Fluoranthene-d10

Concen: 0.357 ng

RT: 19.10 min Scan# 1363

Delta R.T. 0.00 min

Lab File: BE100251.D

Acq: 29 Jun 2019 1:05


Tgt Ion:212 Resp: 34447

Ion Ratio Lower Upper

212 100

106 1.6 1.3 1.9



104 1.0 0.8 1.2




Abundance Ion 212.00 (211.70 to 212.70): E

Ion 106.00 (105.70 to 106.70): E

Ion 104.00 (103.70 to 104.70): E









#26
Fluoranthene
Concen: 2.097 ng
RT: 19.13 min Scan# 1368
Delta R.T. 0.00 min
Lab File: BE100251.D
Acq: 29 Jun 2019 1:05

Instrument :
BNA_E
ClientSampleId :
EXT-001-SW079-061719

Tgt Ion: 202 Resp: 53836
Ion Ratio Lower Upper
202 100
101 5.9 4.6 7.0
203 17.7 13.7 20.5




Abundance Ion 202.00 (201.70 to 202.70): E
Ion 101.00 (100.70 to 101.70): E
Ion 203.00 (202.70 to 203.70): E






#27
Chrysene-d12
Concen: 0.400 ng
RT: 21.25 min Scan# 1620
Delta R.T. 0.00 min
Lab File: BE100251.D
Acq: 29 Jun 2019 1:05

Tgt Ion: 240 Resp: 8956
Ion Ratio Lower Upper
240 100
120 4.4 3.3 4.9
236 28.9 22.6 34.0



Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E









#28
Pyrene
Concen: 2.886 ng
RT: 19.49 min Scan# 1431
Delta R.T. 0.00 min
Lab File: BE100251.D
Acq: 29 Jun 2019 1:05




Instrument :
BNA_E
ClientSampleId :
EXT-001-SW079-061719

Tgt Ion: 202 Resp: 68428
Ion Ratio Lower Upper
202 100
200 19.6 15.6 23.4
203 20.0 14.0 21.0



#29
Terphenyl-d14
Concen: 0.330 ng
RT: 19.70 min Scan# 1467
Delta R.T. -0.01 min
Lab File: BE100251.D
Acq: 29 Jun 2019 1:05

Tgt Ion: 244 Resp: 6132
Ion Ratio Lower Upper
244 100
212 35.0 27.3 40.9
122 5.9 4.7 7.1





#30

Benzo(a)anthracene

Concen: 1.124 ng

RT: 21.22 min Scan# 1618

Delta R.T. 0.00 min

Lab File: BE100251.D

Acq: 29 Jun 2019 1:05

Instrument :

BNA_E

ClientSampleId :

EXT-001-SW079-061719


Tgt Ion:228 Resp: 33094

Ion Ratio Lower Upper

228 100

226 30.9 23.2 34.8

229 22.6 15.7 23.5






Abundance

Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#31

Chrysene

Concen: 1.469 ng

RT: 21.28 min Scan# 1623

Delta R.T. 0.00 min

Lab File: BE100251.D

Acq: 29 Jun 2019 1:05


Tgt Ion:228 Resp: 43447

Ion Ratio Lower Upper

228 100

226 31.1 23.9 35.9

229 22.7 16.5 24.7






Abundance

Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E









#32
 Bis(2-ethylhexyl)phthalate
 Concen: 0.547 ng
 RT: 21.15 min Scan# 1612
 Delta R.T. 0.00 min
 Lab File: BE100251.D
 Acq: 29 Jun 2019 1:05




Instrument :
 BNA_E
 ClientSampleId :
 EXT-001-SW079-061719


Tgt Ion:149 Resp: 25167
 Ion Ratio Lower Upper
 149 100
 167 6.6 5.8 8.8
 279 1.6 1.1 1.7



#33
 Indeno(1,2,3-cd)pyrene
 Concen: 0.653 ng
 RT: 26.22 min Scan# 2855
 Delta R.T. 0.01 min
 Lab File: BE100251.D
 Acq: 29 Jun 2019 1:05

Tgt Ion:276 Resp: 25528
 Ion Ratio Lower Upper
 276 100
 138 9.9 9.4 14.0
 277 24.2 19.2 28.8









#34
Perylene-d12
Concen: 0.400 ng
RT: 23.67 min Scan# 2143
Delta R.T. 0.00 min
Lab File: BE100251.D
Acq: 29 Jun 2019 1:05




Instrument :
BNA_E
ClientSampleId :
EXT-001-SW079-061719


Tgt Ion: 264 Resp: 10772
Ion Ratio Lower Upper
264 100
260 27.2 20.8 31.2
265 32.6 21.8 32.6



#35
Benzo(b)fluoranthene
Concen: 1.517 ng
RT: 22.93 min Scan# 1935
Delta R.T. 0.00 min
Lab File: BE100251.D
Acq: 29 Jun 2019 1:05

Tgt Ion: 252 Resp: 44293
Ion Ratio Lower Upper
252 100
253 24.5 19.2 28.8
125 8.9 7.4 11.0





#36

Benzo(k)fluoranthene

Concen: 1.359 ng

RT: 22.93 min Scan# 1935

Delta R.T. -0.04 min

Lab File: BE100251.D

Acq: 29 Jun 2019 1:05

Instrument :

BNA_E

ClientSampleId :

EXT-001-SW079-061719


Tgt Ion:252 Resp: 44293

Ion Ratio Lower Upper

252 100

253 24.5 17.9 26.9

125 8.9 5.0 7.4#






Abundance

Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#37

Benzo(a)pyrene

Concen: 1.142 ng

RT: 23.56 min Scan# 2113

Delta R.T. 0.01 min

Lab File: BE100251.D

Acq: 29 Jun 2019 1:05


Tgt Ion:252 Resp: 33427

Ion Ratio Lower Upper

252 100

253 24.5 19.6 29.4

125 8.7 9.0 13.6#






Abundance

Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#38

Dibenzo(a,h)anthracene

Concen: 0.218 ng

RT: 26.23 min Scan# 2858

Delta R.T. -0.01 min

Lab File: BE100251.D

Acq: 29 Jun 2019 1:05

Instrument :

BNA_E

ClientSampleId :

EXT-001-SW079-061719


Tgt Ion:278 Resp: 6655

Ion Ratio Lower Upper

278 100

139 16.8 8.2 12.2#




279 35.9 20.3 30.5#



Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E



#39

Benzo(g,h,i)perylene

Concen: 0.892 ng

RT: 27.02 min Scan# 3078

Delta R.T. 0.01 min

Lab File: BE100251.D

Acq: 29 Jun 2019 1:05


Tgt Ion:276 Resp: 28021

Ion Ratio Lower Upper

276 100

277 25.5 20.2 30.2

138 11.0 9.0 13.6



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

