

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062819\
 Data File : BE100233.D
 Acq On : 28 Jun 2019 14:06
 Operator : JU/SJ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE062819

Quant Time: Jun 28 14:45:19 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 28 14:06:53 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.01
2	1,4-Dioxane	0.621	0.608	2.1	97	0.00
3	n-Nitrosodimethylamine	0.289	0.293	-1.4	126	0.00
4 S	2-Fluorophenol	1.020	1.067	-4.6	107	0.01
5 S	Phenol-d6	1.342	1.312	2.2	100	0.00
6	bis(2-Chloroethyl)ether	1.102	1.198	-8.7	100	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
8 S	Nitrobenzene-d5	0.397	0.411	-3.5	105	0.00
9	Naphthalene	4.411	4.454	-1.0	99	0.01
10	Hexachlorobutadiene	0.425	0.429	-0.9	98	0.01
11 SURR	2-Methylnaphthalene-d10	0.752	0.759	-0.9	99	0.01
12	2-Methylnaphthalene	0.774	0.744	3.9	101	0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	102	0.01
14 S	2,4,6-Tribromophenol	0.280	0.281	-0.4	104	0.00
15 S	2-Fluorobiphenyl	1.559	1.545	0.9	99	0.01
16	Acenaphthylene	1.920	1.939	-1.0	100	0.01
17	Acenaphthene	1.088	1.105	-1.6	98	0.00
18	Fluorene	1.609	1.655	-2.9	102	0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.01
20	4-Bromophenyl-phenylether	0.255	0.251	1.6	95	0.00
21	Hexachlorobenzene	0.265	0.271	-2.3	99	0.01
22	Pentachlorophenol	0.080	0.071	11.3	100	-0.01
23	Phenanthrene	0.969	0.998	-3.0	104	0.00
24	Anthracene	0.873	0.889	-1.8	105	0.00
25 SURR	Fluoranthene-d10	5.805	5.882	-1.3	101	0.00
26	Fluoranthene	1.544	1.585	-2.7	101	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	101	0.00
28	Pvrene	1.059	1.096	-3.5	102	0.00
29 S	Terphenyl-d14	0.829	0.846	-2.1	102	0.00
30	Benzo(a)anthracene	1.315	1.339	-1.8	101	0.01
31	Chrysene	1.321	1.403	-6.2	104	0.00
32	Bis(2-ethylhexyl)phthalate	2.255	2.152	4.6	99	0.01
33	Indeno(1,2,3-cd)pyrene	1.747	1.791	-2.5	96	0.01
34 I	Perylene-d12	1.000	1.000	0.0	98	0.01
35	Benzo(b)fluoranthene	1.084	1.124	-3.7	100	0.00
36	Benzo(k)fluoranthene	1.211	1.269	-4.8	99	0.00
37 C	Benzo(a)pyrene	1.087	1.144	-5.2	99	0.00
38	Dibenzo(a,h)anthracene	1.135	1.124	1.0	96	0.00
39	Benzo(a,h,i)perylene	1.166	1.176	-0.9	95	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062819\
Data File : BE100233.D
Acq On : 28 Jun 2019 14:06
Operator : JU/SJ
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE062819

Quant Time: Jun 28 14:45:19 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 28 14:06:53 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			