

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062819\
 Data File : BE100238.D
 Acq On : 28 Jun 2019 17:14
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 29 00:27:32 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 28 14:06:53 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	80	0.00
2	1,4-Dioxane	0.621	0.654	-5.3	81	0.00
3	n-Nitrosodimethylamine	0.289	0.318	-10.0	106	0.01
4 S	2-Fluorophenol	1.020	1.013	0.7	79	0.00
5 S	Phenol-d6	1.342	1.151	14.2	69	0.00
6	bis(2-Chloroethyl)ether	1.102	1.209	-9.7	79	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	75	0.00
8 S	Nitrobenzene-d5	0.397	0.399	-0.5	76	0.00
9	Naphthalene	4.411	4.638	-5.1	77	0.00
10	Hexachlorobutadiene	0.425	0.489	-15.1	83	0.00
11 SURR	2-Methylnaphthalene-d10	0.752	0.803	-6.8	78	0.00
12	2-Methylnaphthalene	0.774	0.751	3.0	76	0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	79	0.01
14 S	2,4,6-Tribromophenol	0.280	0.261	6.8	75	0.00
15 S	2-Fluorobiphenyl	1.559	1.617	-3.7	80	0.01
16	Acenaphthylene	1.920	2.043	-6.4	82	0.00
17	Acenaphthene	1.088	1.136	-4.4	79	0.00
18	Fluorene	1.609	1.681	-4.5	80	0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	77	0.01
20	4-Bromophenyl-phenylether	0.255	0.276	-8.2	80	0.00
21	Hexachlorobenzene	0.265	0.291	-9.8	81	0.00
22	Pentachlorophenol	0.080	0.055	31.3#	59	0.01
23	Phenanthrene	0.969	1.005	-3.7	79	0.00
24	Anthracene	0.873	0.892	-2.2	80	0.00
25 SURR	Fluoranthene-d10	5.805	6.007	-3.5	79	0.00
26	Fluoranthene	1.544	1.641	-6.3	80	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	80	0.00
28	Pvrene	1.059	1.047	1.1	77	0.00
29 S	Terphenyl-d14	0.829	0.815	1.7	78	0.00
30	Benzo(a)anthracene	1.315	1.296	1.4	77	0.01
31	Chrysene	1.321	1.413	-7.0	83	0.00
32	Bis(2-ethylhexyl)phthalate	2.255	1.649	26.9#	60	0.00
33	Indeno(1,2,3-cd)pyrene	1.747	1.803	-3.2	76	0.00
34 I	Perylene-d12	1.000	1.000	0.0	74	0.00
35	Benzo(b)fluoranthene	1.084	1.122	-3.5	76	0.00
36	Benzo(k)fluoranthene	1.211	1.291	-6.6	76	0.00
37 C	Benzo(a)pyrene	1.087	1.194	-9.8	78	0.00
38	Dibenzo(a,h)anthracene	1.135	1.186	-4.5	77	0.00
39	Benzo(a,h,i)perylene	1.166	1.260	-8.1	77	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062819\
Data File : BE100238.D
Acq On : 28 Jun 2019 17:14
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 29 00:27:32 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 28 14:06:53 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			