

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062819\
 Data File : BE100257.D
 Acq On : 29 Jun 2019 4:40
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Jun 29 05:16:25 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jun 29 00:22:29 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	200#	0.00
2	1,4-Dioxane	0.621	0.419	32.5#	130	0.00
3	n-Nitrosodimethylamine	0.289	0.294	-1.7	246#	-0.01
4 S	2-Fluorophenol	1.020	1.147	-12.5	223#	0.00
5 S	Phenol-d6	1.342	1.552	-15.6	232#	-0.01
6	bis(2-Chloroethyl)ether	1.102	1.122	-1.8	183#	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	199#	-0.01
8 S	Nitrobenzene-d5	0.397	0.419	-5.5	213#	-0.01
9	Naphthalene	4.411	4.414	-0.1	195#	0.00
10	Hexachlorobutadiene	0.425	0.419	1.4	190#	0.00
11 SURR	2-Methylnaphthalene-d10	0.752	0.789	-4.9	205#	0.00
12	2-Methylnaphthalene	0.774	0.774	0.0	208#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	199#	0.02
14 S	2,4,6-Tribromophenol	0.280	0.300	-7.1	217#	-0.01
15 S	2-Fluorobiphenyl	1.559	1.560	-0.1	193#	0.00
16	Acenaphthylene	1.920	2.594	-35.1#	259#	0.00
17	Acenaphthene	1.088	1.098	-0.9	190#	0.00
18	Fluorene	1.609	1.589	1.2	190#	0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	167#	0.01
20	4-Bromophenyl-phenylether	0.255	0.285	-11.8	178#	-0.01
21	Hexachlorobenzene	0.265	0.281	-6.0	169#	0.01
22	Pentachlorophenol	0.080	0.110	-37.5#	256#	-0.03
23	Phenanthrene	0.969	1.005	-3.7	172#	0.00
24	Anthracene	0.873	1.017	-16.5	196#	0.00
25 SURR	Fluoranthene-d10	5.805	5.725	1.4	162#	0.00
26	Fluoranthene	1.544	1.478	4.3	155#	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	135	0.00
28	Pvrene	1.059	1.250	-18.0	157#	0.00
29 S	Terphenyl-d14	0.829	0.976	-17.7	158#	0.00
30	Benzo(a)anthracene	1.315	1.471	-11.9	148	0.01
31	Chrysene	1.321	1.347	-2.0	135	0.00
32	Bis(2-ethylhexyl)phthalate	2.255	2.245	0.4	139	0.01
33	Indeno(1,2,3-cd)pyrene	1.747	1.603	8.2	115	0.02
34 I	Perylene-d12	1.000	1.000	0.0	127	0.02
35	Benzo(b)fluoranthene	1.084	1.187	-9.5	137	0.00
36	Benzo(k)fluoranthene	1.211	1.151	5.0	116	0.01
37 C	Benzo(a)pyrene	1.087	1.114	-2.5	125	0.02
38	Dibenzo(a,h)anthracene	1.135	1.117	1.6	124	0.01
39	Benzo(a,h,i)perylene	1.166	1.081	7.3	113	0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062819\
Data File : BE100257.D
Acq On : 29 Jun 2019 4:40
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 29 05:16:25 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jun 29 00:22:29 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			