

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062819\
 Data File : BE100257.D
 Acq On : 29 Jun 2019 4:40
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Jun 29 05:16:25 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jun 29 00:22:29 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	200	0.00
2	1,4-Dioxane	0.400	0.270	32.5#	130	0.00
3	n-Nitrosodimethylamine	0.400	0.406	-1.5	246	-0.01
4 S	2-Fluorophenol	0.400	0.450	-12.5	223	0.00
5 S	Phenol-d6	0.400	0.463	-15.8	232	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.407	-1.7	183	-0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	199	-0.01
8 S	Nitrobenzene-d5	0.400	0.421	-5.2	213	-0.01
9	Naphthalene	0.400	0.400	0.0	195	0.00
10	Hexachlorobutadiene	0.400	0.394	1.5	190	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.420	-5.0	205	0.00
12	2-Methylnaphthalene	0.400	0.400	0.0	208	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	199	0.02
14 S	2,4,6-Tribromophenol	0.400	0.429	-7.2	217	-0.01
15 S	2-Fluorobiphenyl	0.400	0.400	0.0	193	0.00
16	Acenaphthylene	0.400	0.541	-35.3#	259	0.00
17	Acenaphthene	0.400	0.404	-1.0	190	0.00
18	Fluorene	0.400	0.395	1.3	190	0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	167	0.01
20	4-Bromophenyl-phenylether	0.400	0.446	-11.5	178	-0.01
21	Hexachlorobenzene	0.400	0.425	-6.2	169	0.01
22	Pentachlorophenol	0.400	0.550	-37.5#	256	-0.03
23	Phenanthrene	0.400	0.415	-3.7	172	0.00
24	Anthracene	0.400	0.466	-16.5	196	0.00
25 SURR	Fluoranthene-d10	0.400	0.394	1.5	162	0.00
26	Fluoranthene	0.400	0.383	4.3	155	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	135	0.00
28	Pvrene	0.400	0.472	-18.0	157	0.00
29 S	Terphenyl-d14	0.400	0.471	-17.7	158	0.00
30	Benzo(a)anthracene	0.400	0.447	-11.7	148	0.01
31	Chrysene	0.400	0.408	-2.0	135	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.426	-6.5	139	0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.367	8.3	115	0.02
34 I	Perylene-d12	0.400	0.400	0.0	127	0.02
35	Benzo(b)fluoranthene	0.400	0.438	-9.5	137	0.00
36	Benzo(k)fluoranthene	0.400	0.380	5.0	116	0.01
37 C	Benzo(a)pyrene	0.400	0.410	-2.5	125	0.02
38	Dibenzo(a,h)anthracene	0.400	0.394	1.5	124	0.01
39	Benzo(a,h,i)perylene	0.400	0.371	7.3	113	0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062819\
Data File : BE100257.D
Acq On : 29 Jun 2019 4:40
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 29 05:16:25 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jun 29 00:22:29 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				