

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE070119\
 Data File : BE100259.D
 Acq On : 1 Jul 2019 11:15
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 01 14:50:24 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 28 14:06:53 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	117	0.00
2	1,4-Dioxane	0.621	0.700	-12.7	128	0.00
3	n-Nitrosodimethylamine	0.289	0.429	-48.4#	210#	-0.01
4 S	2-Fluorophenol	1.020	1.276	-25.1#	145	0.00
5 S	Phenol-d6	1.342	1.555	-15.9	136	-0.01
6	bis(2-Chloroethyl)ether	1.102	1.175	-6.6	112	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	109	-0.01
8 S	Nitrobenzene-d5	0.397	0.441	-11.1	122	-0.01
9	Naphthalene	4.411	4.633	-5.0	112	0.00
10	Hexachlorobutadiene	0.425	0.449	-5.6	111	0.00
11 SURR	2-Methylnaphthalene-d10	0.752	0.798	-6.1	113	-0.01
12	2-Methylnaphthalene	0.774	0.793	-2.5	117	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	117	0.00
14 S	2,4,6-Tribromophenol	0.280	0.426	-52.1#	180#	-0.01
15 S	2-Fluorobiphenyl	1.559	1.636	-4.9	119	0.00
16	Acenaphthylene	1.920	2.073	-8.0	121	0.00
17	Acenaphthene	1.088	1.191	-9.5	121	0.00
18	Fluorene	1.609	1.781	-10.7	125	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	125	0.01
20	4-Bromophenyl-phenylether	0.255	0.269	-5.5	126	-0.01
21	Hexachlorobenzene	0.265	0.264	0.4	119	0.01
22	Pentachlorophenol	0.080	0.066	17.5	115	-0.01
23	Phenanthrene	0.969	1.025	-5.8	131	0.00
24	Anthracene	0.873	0.999	-14.4	144	0.00
25 SURR	Fluoranthene-d10	5.805	6.070	-4.6	129	0.00
26	Fluoranthene	1.544	1.615	-4.6	127	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	116	0.00
28	Pvrene	1.059	1.199	-13.2	129	0.00
29 S	Terphenyl-d14	0.829	0.963	-16.2	134	0.00
30	Benzo(a)anthracene	1.315	1.432	-8.9	124	0.01
31	Chrysene	1.321	1.456	-10.2	125	0.00
32	Bis(2-ethylhexyl)phthalate	2.255	2.196	2.6	116	0.00
33	Indeno(1,2,3-cd)pyrene	1.747	1.670	4.4	103	0.02
34 I	Perylene-d12	1.000	1.000	0.0	105	0.01
35	Benzo(b)fluoranthene	1.084	1.172	-8.1	112	0.00
36	Benzo(k)fluoranthene	1.211	1.268	-4.7	106	0.00
37 C	Benzo(a)pyrene	1.087	1.164	-7.1	108	0.01
38	Dibenzo(a,h)anthracene	1.135	1.178	-3.8	108	0.02
39	Benzo(a,h,i)perylene	1.166	1.196	-2.6	103	0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE070119\
Data File : BE100259.D
Acq On : 1 Jul 2019 11:15
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 01 14:50:24 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 28 14:06:53 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			