

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE070119\
 Data File : BE100276.D
 Acq On : 1 Jul 2019 21:30
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Jul 02 02:07:11 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 28 14:06:53 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	252#	0.00
2	1,4-Dioxane	0.621	0.401	35.4#	157#	0.01
3	n-Nitrosodimethylamine	0.289	0.393	-36.0#	416#	-0.02
4 S	2-Fluorophenol	1.020	0.952	6.7	234#	0.00
5 S	Phenol-d6	1.342	1.766	-31.6#	333#	0.00
6	bis(2-Chloroethyl)ether	1.102	1.642	-49.0#	339#	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	343#	-0.01
8 S	Nitrobenzene-d5	0.397	0.424	-6.8	372#	-0.03
9	Naphthalene	4.411	4.476	-1.5	342#	0.00
10	Hexachlorobutadiene	0.425	0.432	-1.6	338#	0.00
11 SURR	2-Methylnaphthalene-d10	0.752	0.822	-9.3	369#	-0.01
12	2-Methylnaphthalene	0.774	0.804	-3.9	373#	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	365#	0.00
14 S	2,4,6-Tribromophenol	0.280	0.309	-10.4	409#	-0.01
15 S	2-Fluorobiphenyl	1.559	1.610	-3.3	366#	-0.01
16	Acenaphthylene	1.920	2.000	-4.2	367#	0.00
17	Acenaphthene	1.088	1.116	-2.6	354#	0.00
18	Fluorene	1.609	1.618	-0.6	354#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	318#	0.00
20	4-Bromophenyl-phenylether	0.255	0.280	-9.8	335#	-0.01
21	Hexachlorobenzene	0.265	0.279	-5.3	321#	0.01
22	Pentachlorophenol	0.080	0.101	-26.3#	452#	-0.04
23	Phenanthrene	0.969	1.025	-5.8	335#	-0.01
24	Anthracene	0.873	0.971	-11.2	358#	-0.01
25 SURR	Fluoranthene-d10	5.805	5.711	1.6	309#	0.00
26	Fluoranthene	1.544	1.510	2.2	303#	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	261#	0.00
28	Pvrene	1.059	1.278	-20.7	308#	0.00
29 S	Terphenyl-d14	0.829	1.009	-21.7	314#	0.00
30	Benzo(a)anthracene	1.315	1.462	-11.2	284#	0.01
31	Chrysene	1.321	1.439	-8.9	277#	0.00
32	Bis(2-ethylhexyl)phthalate	2.255	2.229	1.2	265#	0.00
33	Indeno(1,2,3-cd)pyrene	1.747	1.749	-0.1	241#	0.01
34 I	Perylene-d12	1.000	1.000	0.0	240#	0.00
35	Benzo(b)fluoranthene	1.084	1.163	-7.3	254#	0.00
36	Benzo(k)fluoranthene	1.211	1.216	-0.4	232#	0.00
37 C	Benzo(a)pyrene	1.087	1.134	-4.3	241#	0.00
38	Dibenzo(a,h)anthracene	1.135	1.202	-5.9	252#	0.01
39	Benzo(a,h,i)perylene	1.166	1.222	-4.8	242#	0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE070119\
Data File : BE100276.D
Acq On : 1 Jul 2019 21:30
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 02 02:07:11 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 28 14:06:53 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			