

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE070119\
 Data File : BE100278.D
 Acq On : 1 Jul 2019 23:56
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 02 02:01:51 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 28 14:06:53 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	164#	0.00
2	1,4-Dioxane	0.621	0.497	20.0	128	0.01
3	n-Nitrosodimethylamine	0.289	0.312	-8.0	215#	0.00
4 S	2-Fluorophenol	1.020	1.091	-7.0	175#	0.01
5 S	Phenol-d6	1.342	1.470	-9.5	181#	-0.01
6	bis(2-Chloroethyl)ether	1.102	1.233	-11.9	166#	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	160#	-0.01
8 S	Nitrobenzene-d5	0.397	0.456	-14.9	186#	-0.01
9	Naphthalene	4.411	4.622	-4.8	164#	0.00
10	Hexachlorobutadiene	0.425	0.471	-10.8	172#	0.00
11 SURR	2-Methylnaphthalene-d10	0.752	0.816	-8.5	170#	-0.01
12	2-Methylnaphthalene	0.774	0.799	-3.2	173#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	161#	0.00
14 S	2,4,6-Tribromophenol	0.280	0.276	1.4	161#	0.00
15 S	2-Fluorobiphenyl	1.559	1.666	-6.9	167#	0.00
16	Acenaphthylene	1.920	2.056	-7.1	166#	0.00
17	Acenaphthene	1.088	1.165	-7.1	163#	0.00
18	Fluorene	1.609	1.617	-0.5	156#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	137	0.01
20	4-Bromophenyl-phenylether	0.255	0.286	-12.2	147	0.00
21	Hexachlorobenzene	0.265	0.301	-13.6	149	0.01
22	Pentachlorophenol	0.080	0.061	23.8	118	-0.03
23	Phenanthrene	0.969	1.043	-7.6	147	0.00
24	Anthracene	0.873	0.958	-9.7	153#	0.00
25 SURR	Fluoranthene-d10	5.805	5.603	3.5	131	0.00
26	Fluoranthene	1.544	1.537	0.5	133	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	112	0.00
28	Pvrene	1.059	1.303	-23.0	135	0.00
29 S	Terphenyl-d14	0.829	1.024	-23.5	137	0.00
30	Benzo(a)anthracene	1.315	1.441	-9.6	120	0.01
31	Chrysene	1.321	1.520	-15.1	125	0.00
32	Bis(2-ethylhexyl)phthalate	2.255	2.071	8.2	105	0.00
33	Indeno(1,2,3-cd)pyrene	1.747	1.772	-1.4	105	0.02
34 I	Perylene-d12	1.000	1.000	0.0	101	0.01
35	Benzo(b)fluoranthene	1.084	1.184	-9.2	109	0.00
36	Benzo(k)fluoranthene	1.211	1.280	-5.7	103	0.00
37 C	Benzo(a)pyrene	1.087	1.174	-8.0	105	0.01
38	Dibenzo(a,h)anthracene	1.135	1.215	-7.0	108	0.01
39	Benzo(a,h,i)perylene	1.166	1.249	-7.1	104	0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE070119\
Data File : BE100278.D
Acq On : 1 Jul 2019 23:56
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 02 02:01:51 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 28 14:06:53 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			