

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE070119\
 Data File : BE100291.D
 Acq On : 2 Jul 2019 7:45
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 02 08:28:53 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 28 14:06:53 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	340#	0.00
2	1,4-Dioxane	0.621	0.387	37.7#	205#	0.01
3	n-Nitrosodimethylamine	0.289	0.366	-26.6#	522#	-0.01
4 S	2-Fluorophenol	1.020	1.013	0.7	336#	0.01
5 S	Phenol-d6	1.342	1.756	-30.8#	447#	-0.01
6	bis(2-Chloroethyl)ether	1.102	1.618	-46.8#	451#	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	456#	-0.01
8 S	Nitrobenzene-d5	0.397	0.446	-12.3	521#	-0.01
9	Naphthalene	4.411	4.482	-1.6	455#	0.00
10	Hexachlorobutadiene	0.425	0.426	-0.2	443#	0.00
11 SURR	2-Methylnaphthalene-d10	0.752	0.800	-6.4	477#	-0.01
12	2-Methylnaphthalene	0.774	0.781	-0.9	482#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	459#	0.00
14 S	2,4,6-Tribromophenol	0.280	0.315	-12.5	524#	-0.01
15 S	2-Fluorobiphenyl	1.559	1.601	-2.7	457#	-0.01
16	Acenaphthylene	1.920	1.932	-0.6	445#	0.00
17	Acenaphthene	1.088	1.129	-3.8	450#	0.00
18	Fluorene	1.609	1.589	1.2	437#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	399#	0.01
20	4-Bromophenyl-phenylether	0.255	0.279	-9.4	418#	-0.01
21	Hexachlorobenzene	0.265	0.283	-6.8	407#	0.01
22	Pentachlorophenol	0.080	0.100	-25.0#	557#	-0.03
23	Phenanthrene	0.969	1.022	-5.5	418#	0.00
24	Anthracene	0.873	1.005	-15.1	465#	0.00
25 SURR	Fluoranthene-d10	5.805	5.733	1.2	389#	0.00
26	Fluoranthene	1.544	1.500	2.8	377#	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	325#	0.00
28	Pvrene	1.059	1.272	-20.1	382#	0.00
29 S	Terphenyl-d14	0.829	0.988	-19.2	383#	0.00
30	Benzo(a)anthracene	1.315	1.463	-11.3	354#	0.01
31	Chrysene	1.321	1.436	-8.7	344#	0.00
32	Bis(2-ethylhexyl)phthalate	2.255	2.284	-1.3	338#	0.00
33	Indeno(1,2,3-cd)pyrene	1.747	1.612	7.7	277#	0.03
34 I	Perylene-d12	1.000	1.000	0.0	307#	0.02
35	Benzo(b)fluoranthene	1.084	1.161	-7.1	324#	0.00
36	Benzo(k)fluoranthene	1.211	1.206	0.4	294#	0.01
37 C	Benzo(a)pyrene	1.087	1.144	-5.2	310#	0.01
38	Dibenzo(a,h)anthracene	1.135	1.127	0.7	302#	0.02
39	Benzo(a,h,i)perylene	1.166	1.020	12.5	258#	0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE070119\
Data File : BE100291.D
Acq On : 2 Jul 2019 7:45
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 02 08:28:53 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 28 14:06:53 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			