


Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE070119\
 Data File : BE100258.D
 Acq On : 1 Jul 2019 10:40
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE062819.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Jun 28 14:06:53 2019



AutoFind: Scans 2537, 2538, 2539; Background Corrected with Scan 2529

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	21.8	2896	PASS
68	69	0.00	2	1.4	60	PASS
69	198	0.00	100	32.8	4353	PASS
70	69	0.00	2	0.5	20	PASS
127	198	10	80	33.8	4491	PASS
197	198	0.00	2	0.2	23	PASS
198	198	100	100	100.0	13290	PASS
199	198	5	9	7.0	926	PASS
275	198	10	60	30.8	4098	PASS
365	198	1	100	4.0	538	PASS
441	198	0.01	100	13.3	1770	PASS
442	198	50	100	83.7	11124	PASS
443	442	15	24	19.6	2177	PASS