

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE070219\
 Data File : BE100298.D
 Acq On : 2 Jul 2019 12:35
 Operator : JU/SJ
 Sample : PB120959BL
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 PB120959BL

Quant Time: Jul 03 03:13:27 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 28 14:06:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.65	152	778	0.40	ng	0.00
7) Naphthalene-d8	10.45	136	2341	0.40	ng	0.00
13) Acenaphthene-d10	14.33	164	1836	0.40	ng	0.01
19) Phenanthrene-d10	17.07	188	5916	0.40	ng	0.01
27) Chrysene-d12	21.26	240	9289	0.40	ng	0.01
34) Perylene-d12	23.68	264	11887	0.40	ng	0.02

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.22	112	690	0.35	ng	0.00
5) Phenol-d6	6.85	99	706	0.27	ng	0.01
8) Nitrobenzene-d5	8.85	82	544	0.23	ng	0.03
11) 2-Methylnaphthalene-d10	12.06	152	1305	0.30	ng	0.01
14) 2,4,6-Tribromophenol	15.84	330	344	0.27	ng	0.01
15) 2-Fluorobiphenyl	12.96	172	1998	0.28	ng	0.01
25) Fluoranthene-d10	19.10	212	27482	0.32	ng	0.00
29) Terphenyl-d14	19.71	244	5993	0.31	ng	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE070219\
Data File : BE100298.D
Acq On : 2 Jul 2019 12:35
Operator : JU/SJ
Sample : PB120959BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
PB120959BL

Quant Time: Jul 03 03:13:27 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 28 14:06:53 2019
Response via : Initial Calibration

