

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE070219\
 Data File : BE100293.D
 Acq On : 2 Jul 2019 9:35
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Jul 03 02:32:27 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 28 14:06:53 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	1,4-Dioxane	0.621	0.568	8.5	89	0.00
3	n-Nitrosodimethylamine	0.289	0.328	-13.5	139	-0.01
4 S	2-Fluorophenol	1.020	1.364	-33.7#	134	0.00
5 S	Phenol-d6	1.342	1.481	-10.4	112	0.00
6	bis(2-Chloroethyl)ether	1.102	1.168	-6.0	96	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	91	-0.01
8 S	Nitrobenzene-d5	0.397	0.439	-10.6	103	-0.01
9	Naphthalene	4.411	4.627	-4.9	94	0.00
10	Hexachlorobutadiene	0.425	0.427	-0.5	89	0.00
11 SURR	2-Methylnaphthalene-d10	0.752	0.765	-1.7	91	0.00
12	2-Methylnaphthalene	0.774	0.755	2.5	93	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	93	0.02
14 S	2,4,6-Tribromophenol	0.280	0.429	-53.2#	145	0.00
15 S	2-Fluorobiphenyl	1.559	1.631	-4.6	95	0.00
16	Acenaphthylene	1.920	2.141	-11.5	100	0.00
17	Acenaphthene	1.088	1.161	-6.7	94	0.00
18	Fluorene	1.609	1.747	-8.6	98	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	92	0.01
20	4-Bromophenyl-phenylether	0.255	0.276	-8.2	95	-0.01
21	Hexachlorobenzene	0.265	0.283	-6.8	94	0.01
22	Pentachlorophenol	0.080	0.025	68.8#	32#	0.01
23	Phenanthrene	0.969	1.066	-10.0	101	0.00
24	Anthracene	0.873	1.026	-17.5	110	0.00
25 SURR	Fluoranthene-d10	5.805	6.284	-8.3	98	0.00
26	Fluoranthene	1.544	1.777	-15.1	103	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	89	0.00
28	Pvrene	1.059	1.300	-22.8	107	0.00
29 S	Terphenyl-d14	0.829	0.916	-10.5	97	0.00
30	Benzo(a)anthracene	1.315	1.426	-8.4	95	0.01
31	Chrysene	1.321	1.502	-13.7	99	0.00
32	Bis(2-ethylhexyl)phthalate	2.255	2.413	-7.0	98	0.00
33	Indeno(1,2,3-cd)pyrene	1.747	1.789	-2.4	84	0.02
34 I	Perylene-d12	1.000	1.000	0.0	85	0.01
35	Benzo(b)fluoranthene	1.084	1.196	-10.3	93	0.00
36	Benzo(k)fluoranthene	1.211	1.218	-0.6	82	0.01
37 C	Benzo(a)pyrene	1.087	1.178	-8.4	89	0.02
38	Dibenzo(a,h)anthracene	1.135	1.166	-2.7	87	0.02
39	Benzo(a,h,i)perylene	1.166	1.204	-3.3	85	0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE070219\
Data File : BE100293.D
Acq On : 2 Jul 2019 9:35
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 03 02:32:27 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 28 14:06:53 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			