

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE070219\
 Data File : BE100304.D
 Acq On : 2 Jul 2019 16:14
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Jul 03 02:13:02 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE062819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 28 14:06:53 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
2	1,4-Dioxane	0.621	0.609	1.9	91	0.00
3	n-Nitrosodimethylamine	0.289	0.262	9.3	105	0.00
4 S	2-Fluorophenol	1.020	1.097	-7.5	102	0.00
5 S	Phenol-d6	1.342	1.258	6.3	90	0.00
6	bis(2-Chloroethyl)ether	1.102	1.095	0.6	86	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	88	-0.01
8 S	Nitrobenzene-d5	0.397	0.400	-0.8	90	0.00
9	Naphthalene	4.411	4.541	-2.9	89	0.00
10	Hexachlorobutadiene	0.425	0.458	-7.8	92	0.00
11 SURR	2-Methylnaphthalene-d10	0.752	0.767	-2.0	88	0.00
12	2-Methylnaphthalene	0.774	0.737	4.8	88	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	92	0.00
14 S	2,4,6-Tribromophenol	0.280	0.298	-6.4	100	0.01
15 S	2-Fluorobiphenyl	1.559	1.566	-0.4	90	0.00
16	Acenaphthylene	1.920	2.006	-4.5	93	0.00
17	Acenaphthene	1.088	1.117	-2.7	90	0.00
18	Fluorene	1.609	1.680	-4.4	93	0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	89	0.01
20	4-Bromophenyl-phenylether	0.255	0.276	-8.2	92	0.00
21	Hexachlorobenzene	0.265	0.289	-9.1	93	0.01
22	Pentachlorophenol	0.080	0.005	93.8#	7#	0.04
23	Phenanthrene	0.969	1.023	-5.6	93	0.00
24	Anthracene	0.873	0.945	-8.2	97	0.00
25 SURR	Fluoranthene-d10	5.805	6.135	-5.7	93	0.00
26	Fluoranthene	1.544	1.638	-6.1	92	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	81	0.00
28	Pyrene	1.059	1.240	-17.1	93	0.00
29 S	Terphenyl-d14	0.829	0.952	-14.8	93	0.00
30	Benzo(a)anthracene	1.315	1.381	-5.0	84	0.01
31	Chrysene	1.321	1.467	-11.1	88	0.00
32	Bis(2-ethylhexyl)phthalate	2.255	2.152	4.6	80	0.00
33	Indeno(1,2,3-cd)pyrene	1.747	1.893	-8.4	82	0.02
34 I	Perylene-d12	1.000	1.000	0.0	81	0.01
35	Benzo(b)fluoranthene	1.084	1.114	-2.8	82	0.00
36	Benzo(k)fluoranthene	1.211	1.248	-3.1	80	0.01
37 C	Benzo(a)pyrene	1.087	1.188	-9.3	85	0.01
38	Dibenzo(a,h)anthracene	1.135	1.202	-5.9	85	0.01
39	Benzo(g,h,i)perylene	1.166	1.232	-5.7	82	0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE070219\
Data File : BE100304.D
Acq On : 2 Jul 2019 16:14
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 03 02:13:02 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE062819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 28 14:06:53 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			