

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE070219\
 Data File : BE100321.D
 Acq On : 3 Jul 2019 2:31
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Jul 03 04:17:10 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 28 14:06:53 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	137	0.00
2	1,4-Dioxane	0.621	0.450	27.5#	96	0.01
3	n-Nitrosodimethylamine	0.289	0.281	2.8	162#	0.00
4 S	2-Fluorophenol	1.020	1.098	-7.6	147	0.01
5 S	Phenol-d6	1.342	1.559	-16.2	160#	0.00
6	bis(2-Chloroethyl)ether	1.102	1.099	0.3	123	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	133	-0.01
8 S	Nitrobenzene-d5	0.397	0.426	-7.3	145	-0.01
9	Naphthalene	4.411	4.633	-5.0	138	0.00
10	Hexachlorobutadiene	0.425	0.472	-11.1	143	0.00
11 SURR	2-Methylnaphthalene-d10	0.752	0.773	-2.8	135	0.00
12	2-Methylnaphthalene	0.774	0.768	0.8	139	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	128	0.00
14 S	2,4,6-Tribromophenol	0.280	0.308	-10.0	143	-0.01
15 S	2-Fluorobiphenyl	1.559	1.680	-7.8	134	0.00
16	Acenaphthylene	1.920	2.057	-7.1	132	0.00
17	Acenaphthene	1.088	1.156	-6.2	128	0.00
18	Fluorene	1.609	1.645	-2.2	126	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	114	0.01
20	4-Bromophenyl-phenylether	0.255	0.278	-9.0	119	-0.01
21	Hexachlorobenzene	0.265	0.291	-9.8	119	0.01
22	Pentachlorophenol	0.080	0.087	-8.7	139	-0.03
23	Phenanthrene	0.969	1.004	-3.6	117	0.00
24	Anthracene	0.873	0.971	-11.2	128	0.00
25 SURR	Fluoranthene-d10	5.805	5.692	1.9	110	0.00
26	Fluoranthene	1.544	1.520	1.6	109	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	93	0.00
28	Pvrene	1.059	1.276	-20.5	110	0.00
29 S	Terphenyl-d14	0.829	0.980	-18.2	109	0.00
30	Benzo(a)anthracene	1.315	1.400	-6.5	97	0.01
31	Chrysene	1.321	1.415	-7.1	97	0.00
32	Bis(2-ethylhexyl)phthalate	2.255	2.238	0.8	95	0.00
33	Indeno(1,2,3-cd)pyrene	1.747	1.742	0.3	86	0.03
34 I	Perylene-d12	1.000	1.000	0.0	85	0.01
35	Benzo(b)fluoranthene	1.084	1.188	-9.6	92	0.00
36	Benzo(k)fluoranthene	1.211	1.232	-1.7	83	0.00
37 C	Benzo(a)pyrene	1.087	1.174	-8.0	88	0.01
38	Dibenzo(a,h)anthracene	1.135	1.211	-6.7	90	0.02
39	Benzo(a,h,i)perylene	1.166	1.194	-2.4	84	0.03

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE070219\
Data File : BE100321.D
Acq On : 3 Jul 2019 2:31
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 03 04:17:10 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 28 14:06:53 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			