

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE070319\
 Data File : BE100331.D
 Acq On : 3 Jul 2019 16:34
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 04 01:20:03 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 28 14:06:53 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	120	-0.01
2	1,4-Dioxane	0.400	0.351	12.3	102	-0.01
3	n-Nitrosodimethylamine	0.400	0.434	-8.5	158	-0.02
4 S	2-Fluorophenol	0.400	0.358	10.5	107	-0.01
5 S	Phenol-d6	0.400	0.385	3.8	116	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.443	-10.7	120	-0.02
7 I	Naphthalene-d8	0.400	0.400	0.0	118	-0.01
8 S	Nitrobenzene-d5	0.400	0.400	0.0	120	-0.01
9	Naphthalene	0.400	0.420	-5.0	122	0.00
10	Hexachlorobutadiene	0.400	0.469	-17.2	134	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.432	-8.0	126	-0.01
12	2-Methylnaphthalene	0.400	0.391	2.3	121	-0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	125	0.00
14 S	2,4,6-Tribromophenol	0.400	0.344	14.0	109	0.00
15 S	2-Fluorobiphenyl	0.400	0.409	-2.2	124	0.00
16	Acenaphthylene	0.400	0.426	-6.5	128	0.00
17	Acenaphthene	0.400	0.426	-6.5	126	-0.01
18	Fluorene	0.400	0.422	-5.5	127	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	120	0.00
20	4-Bromophenyl-phenylether	0.400	0.442	-10.5	127	-0.01
21	Hexachlorobenzene	0.400	0.444	-11.0	127	0.00
22	Pentachlorophenol	0.400	0.193	51.7#	65	-0.03
23	Phenanthrene	0.400	0.409	-2.2	122	-0.01
24	Anthracene	0.400	0.396	1.0	120	-0.01
25 SURR	Fluoranthene-d10	0.400	0.394	1.5	116	0.00
26	Fluoranthene	0.400	0.410	-2.5	119	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	93	0.00
28	Pvrene	0.400	0.533	-33.3#	122	0.00
29 S	Terphenyl-d14	0.400	0.558	-39.5#	129	0.00
30	Benzo(a)anthracene	0.400	0.420	-5.0	96	0.01
31	Chrysene	0.400	0.433	-8.2	98	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.458	-14.5	102	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.388	3.0	83	0.01
34 I	Perylene-d12	0.400	0.400	0.0	87	0.00
35	Benzo(b)fluoranthene	0.400	0.409	-2.2	88	0.00
36	Benzo(k)fluoranthene	0.400	0.417	-4.2	88	0.00
37 C	Benzo(a)pyrene	0.400	0.423	-5.7	89	0.00
38	Dibenzo(a,h)anthracene	0.400	0.374	6.5	81	0.01
39	Benzo(a,h,i)perylene	0.400	0.429	-7.2	90	0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE070319\
Data File : BE100331.D
Acq On : 3 Jul 2019 16:34
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 04 01:20:03 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 28 14:06:53 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				