

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE070319\
 Data File : BE100331.D
 Acq On : 3 Jul 2019 16:34
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 04 01:20:03 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 28 14:06:53 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	120	-0.01
2	1,4-Dioxane	0.621	0.544	12.4	102	-0.01
3	n-Nitrosodimethylamine	0.289	0.314	-8.7	158#	-0.02
4 S	2-Fluorophenol	1.020	0.914	10.4	107	-0.01
5 S	Phenol-d6	1.342	1.291	3.8	116	-0.01
6	bis(2-Chloroethyl)ether	1.102	1.222	-10.9	120	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	118	-0.01
8 S	Nitrobenzene-d5	0.397	0.398	-0.3	120	-0.01
9	Naphthalene	4.411	4.635	-5.1	122	0.00
10	Hexachlorobutadiene	0.425	0.498	-17.2	134	-0.01
11 SURR	2-Methylnaphthalene-d10	0.752	0.813	-8.1	126	-0.01
12	2-Methylnaphthalene	0.774	0.756	2.3	121	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	125	0.00
14 S	2,4,6-Tribromophenol	0.280	0.241	13.9	109	0.00
15 S	2-Fluorobiphenyl	1.559	1.595	-2.3	124	0.00
16	Acenaphthylene	1.920	2.043	-6.4	128	0.00
17	Acenaphthene	1.088	1.158	-6.4	126	-0.01
18	Fluorene	1.609	1.697	-5.5	127	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	120	0.00
20	4-Bromophenyl-phenylether	0.255	0.282	-10.6	127	-0.01
21	Hexachlorobenzene	0.265	0.294	-10.9	127	0.00
22	Pentachlorophenol	0.080	0.039	51.2#	65	-0.03
23	Phenanthrene	0.969	0.992	-2.4	122	-0.01
24	Anthracene	0.873	0.863	1.1	120	-0.01
25 SURR	Fluoranthene-d10	5.805	5.713	1.6	116	0.00
26	Fluoranthene	1.544	1.584	-2.6	119	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	93	0.00
28	Pvrene	1.059	1.410	-33.1#	122	0.00
29 S	Terphenyl-d14	0.829	1.156	-39.4#	129	0.00
30	Benzo(a)anthracene	1.315	1.380	-4.9	96	0.01
31	Chrysene	1.321	1.428	-8.1	98	0.00
32	Bis(2-ethylhexyl)phthalate	2.255	2.394	-6.2	102	0.00
33	Indeno(1,2,3-cd)pyrene	1.747	1.694	3.0	83	0.01
34 I	Perylene-d12	1.000	1.000	0.0	87	0.00
35	Benzo(b)fluoranthene	1.084	1.108	-2.2	88	0.00
36	Benzo(k)fluoranthene	1.211	1.263	-4.3	88	0.00
37 C	Benzo(a)pyrene	1.087	1.148	-5.6	89	0.00
38	Dibenzo(a,h)anthracene	1.135	1.062	6.4	81	0.01
39	Benzo(a,h,i)perylene	1.166	1.250	-7.2	90	0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE070319\
Data File : BE100331.D
Acq On : 3 Jul 2019 16:34
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 04 01:20:03 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 28 14:06:53 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			