

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE070319\
 Data File : BE100333.D
 Acq On : 3 Jul 2019 17:50
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Jul 03 20:55:16 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 28 14:06:53 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	-0.01
2	1,4-Dioxane	0.621	0.627	-1.0	94	-0.01
3	n-Nitrosodimethylamine	0.289	0.271	6.2	109	-0.01
4 S	2-Fluorophenol	1.020	0.902	11.6	85	-0.01
5 S	Phenol-d6	1.342	1.137	15.3	82	0.00
6	bis(2-Chloroethyl)ether	1.102	1.070	2.9	84	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	95	-0.01
8 S	Nitrobenzene-d5	0.397	0.394	0.8	96	-0.01
9	Naphthalene	4.411	4.711	-6.8	100	0.00
10	Hexachlorobutadiene	0.425	0.521	-22.6	113	-0.01
11 SURR	2-Methylnaphthalene-d10	0.752	0.822	-9.3	102	0.00
12	2-Methylnaphthalene	0.774	0.761	1.7	98	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
14 S	2,4,6-Tribromophenol	0.280	0.191	31.8#	69	0.00
15 S	2-Fluorobiphenyl	1.559	1.680	-7.8	104	0.00
16	Acenaphthylene	1.920	2.133	-11.1	107	0.00
17	Acenaphthene	1.088	1.227	-12.8	106	-0.01
18	Fluorene	1.609	1.790	-11.2	107	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	96	0.00
20	4-Bromophenyl-phenylether	0.255	0.287	-12.5	104	-0.01
21	Hexachlorobenzene	0.265	0.301	-13.6	105	0.00
22	Pentachlorophenol	0.080	0.021	73.8#	28#	0.01
23	Phenanthrene	0.969	1.004	-3.6	99	-0.01
24	Anthracene	0.873	0.859	1.6	96	0.00
25 SURR	Fluoranthene-d10	5.805	5.781	0.4	94	0.00
26	Fluoranthene	1.544	1.632	-5.7	99	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	77	0.00
28	Pvrene	1.059	1.409	-33.1#	101	0.00
29 S	Terphenyl-d14	0.829	1.127	-35.9#	104	0.00
30	Benzo(a)anthracene	1.315	1.275	3.0	73	0.01
31	Chrysene	1.321	1.444	-9.3	82	0.00
32	Bis(2-ethylhexyl)phthalate	2.255	2.425	-7.5	85	0.00
33	Indeno(1,2,3-cd)pyrene	1.747	1.639	6.2	67	0.01
34 I	Perylene-d12	1.000	1.000	0.0	69	0.01
35	Benzo(b)fluoranthene	1.084	1.045	3.6	66	0.00
36	Benzo(k)fluoranthene	1.211	1.422	-17.4	78	0.00
37 C	Benzo(a)pyrene	1.087	1.155	-6.3	70	0.00
38	Dibenzo(a,h)anthracene	1.135	1.027	9.5	62	0.01
39	Benzo(a,h,i)perylene	1.166	1.294	-11.0	73	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE070319\
Data File : BE100333.D
Acq On : 3 Jul 2019 17:50
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 03 20:55:16 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 28 14:06:53 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			