

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE070319\
 Data File : BE100350.D
 Acq On : 4 Jul 2019 4:14
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 05 01:42:16 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 28 14:06:53 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	116	-0.01
2	1,4-Dioxane	0.621	0.541	12.9	98	-0.01
3	n-Nitrosodimethylamine	0.289	0.346	-19.7	168#	-0.02
4 S	2-Fluorophenol	1.020	1.028	-0.8	116	-0.01
5 S	Phenol-d6	1.342	1.346	-0.3	117	-0.01
6	bis(2-Chloroethyl)ether	1.102	1.018	7.6	97	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	107	-0.01
8 S	Nitrobenzene-d5	0.397	0.419	-5.5	114	-0.01
9	Naphthalene	4.411	4.719	-7.0	112	-0.01
10	Hexachlorobutadiene	0.425	0.542	-27.5#	132	-0.01
11 SURR	2-Methylnaphthalene-d10	0.752	0.828	-10.1	115	-0.01
12	2-Methylnaphthalene	0.774	0.763	1.4	110	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
14 S	2,4,6-Tribromophenol	0.280	0.313	-11.8	117	-0.01
15 S	2-Fluorobiphenyl	1.559	1.742	-11.7	111	-0.01
16	Acenaphthylene	1.920	2.153	-12.1	111	0.00
17	Acenaphthene	1.088	1.200	-10.3	107	-0.01
18	Fluorene	1.609	1.738	-8.0	107	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
20	4-Bromophenyl-phenylether	0.255	0.285	-11.8	103	-0.01
21	Hexachlorobenzene	0.265	0.310	-17.0	108	0.00
22	Pentachlorophenol	0.080	0.082	-2.5	111	-0.03
23	Phenanthrene	0.969	1.038	-7.1	103	-0.01
24	Anthracene	0.873	0.938	-7.4	105	-0.01
25 SURR	Fluoranthene-d10	5.805	6.006	-3.5	99	0.00
26	Fluoranthene	1.544	1.658	-7.4	101	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	81	0.00
28	Pvrene	1.059	1.373	-29.7#	103	0.00
29 S	Terphenyl-d14	0.829	1.052	-26.9#	102	0.00
30	Benzo(a)anthracene	1.315	1.401	-6.5	84	0.01
31	Chrysene	1.321	1.422	-7.6	85	0.00
32	Bis(2-ethylhexyl)phthalate	2.255	2.715	-20.4	100	0.00
33	Indeno(1,2,3-cd)pyrene	1.747	1.728	1.1	74	0.00
34 I	Perylene-d12	1.000	1.000	0.0	73	0.00
35	Benzo(b)fluoranthene	1.084	1.199	-10.6	79	0.00
36	Benzo(k)fluoranthene	1.211	1.285	-6.1	74	0.00
37 C	Benzo(a)pyrene	1.087	1.163	-7.0	75	0.00
38	Dibenzo(a,h)anthracene	1.135	1.190	-4.8	76	0.00
39	Benzo(a,h,i)perylene	1.166	1.291	-10.7	77	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE070319\
Data File : BE100350.D
Acq On : 4 Jul 2019 4:14
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 05 01:42:16 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 28 14:06:53 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			