

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE070618\
 Data File : BE096856.D
 Acq On : 6 Jul 2018 16:53
 Operator : SJ/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SICV50

Quant Time: Jul 06 17:30:38 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE070618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 06 16:55:25 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	113	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	108	0.00
3	Naphthalene	0.967	0.973	-0.6	109	0.00
4 SURR	2-Methylnaphthalene-d10	0.655	0.653	0.3	106	0.00
5	2-Methylnaphthalene	0.659	0.660	-0.2	106	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	105	0.00
7	Acenaphthylene	1.639	1.629	0.6	105	0.00
8 C	Acenaphthene	1.354	1.339	1.1	105	0.00
9	Fluorene	1.397	1.385	0.9	106	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	104	0.00
11	Pentachlorophenol	0.026	0.021	19.2	102	0.00
12	Phenanthrene	1.031	1.021	1.0	104	0.00
13	Anthracene	0.876	0.855	2.4	103	0.00
14 SURR	Fluoranthene-d10	0.996	0.948	4.8	102	0.00
15 C	Fluoranthene	1.099	1.064	3.2	104	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	103	0.00
17	Pyrene	1.313	1.359	-3.5	107	0.00
18	Benzo(a)anthracene	0.943	0.930	1.4	104	0.00
19	Chrysene	1.222	1.222	0.0	104	0.00
20 I	Perylene-d12	1.000	1.000	0.0	96	0.00
21	Benzo(b)fluoranthene	1.092	1.124	-2.9	102	0.00
22	Benzo(k)fluoranthene	1.192	1.195	-0.3	103	0.00
23 C	Benzo(a)pyrene	0.923	0.922	0.1	101	0.00
24	Indeno(1,2,3-cd)pyrene	0.959	0.946	1.4	99	0.00
25	Dibenzo(a,h)anthracene	0.761	0.761	0.0	98	0.00
26	Benzo(g,h,i)perylene	0.930	0.941	-1.2	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0