

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE070618\
 Data File : BE096857.D
 Acq On : 6 Jul 2018 17:29
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.451

Quant Time: Jul 06 18:07:15 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE070618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 06 16:55:25 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	108	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	104	0.00
3	Naphthalene	0.967	0.977	-1.0	105	0.00
4 SURR	2-Methylnaphthalene-d10	0.655	0.655	0.0	102	0.00
5	2-Methylnaphthalene	0.659	0.661	-0.3	102	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	102	0.00
7	Acenaphthylene	1.639	1.640	-0.1	102	0.00
8 C	Acenaphthene	1.354	1.348	0.4	102	0.00
9	Fluorene	1.397	1.397	0.0	104	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	102	0.00
11	Pentachlorophenol	0.026	0.023	11.5	106	0.00
12	Phenanthrene	1.031	1.032	-0.1	103	0.00
13	Anthracene	0.876	0.857	2.2	101	0.00
14 SURR	Fluoranthene-d10	0.996	0.974	2.2	103	0.00
15 C	Fluoranthene	1.099	1.089	0.9	104	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	104	0.00
17	Pyrene	1.313	1.349	-2.7	108	0.00
18	Benzo(a)anthracene	0.943	0.935	0.8	105	0.00
19	Chrysene	1.222	1.222	0.0	105	0.00
20 I	Perylene-d12	1.000	1.000	0.0	98	0.00
21	Benzo(b)fluoranthene	1.092	1.091	0.1	101	0.00
22	Benzo(k)fluoranthene	1.192	1.205	-1.1	106	0.00
23 C	Benzo(a)pyrene	0.923	0.921	0.2	103	0.00
24	Indeno(1,2,3-cd)pyrene	0.959	0.964	-0.5	103	0.00
25	Dibenzo(a,h)anthracene	0.761	0.774	-1.7	103	0.00
26	Benzo(g,h,i)perylene	0.930	0.939	-1.0	104	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0