

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE070618\
 Data File : BE096874.D
 Acq On : 7 Jul 2018 4:05
 Operator : SJ/JU
 Sample : SSTDC00.4EC
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.452

Quant Time: Jul 07 06:49:15 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE070618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 06 16:55:25 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	97	0.00
3	Naphthalene	0.967	0.966	0.1	97	0.00
4 SURR	2-Methylnaphthalene-d10	0.655	0.676	-3.2	99	0.00
5	2-Methylnaphthalene	0.659	0.682	-3.5	99	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
7	Acenaphthylene	1.639	1.803	-10.0	111	0.00
8 C	Acenaphthene	1.354	1.346	0.6	100	0.00
9	Fluorene	1.397	1.437	-2.9	104	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	89	0.00
11	Pentachlorophenol	0.026	0.077	-196.2#	312#	0.00
12	Phenanthrene	1.031	1.023	0.8	89	0.00
13	Anthracene	0.876	1.027	-17.2	106	0.00
14 SURR	Fluoranthene-d10	0.996	1.179	-18.4	108	0.00
15 C	Fluoranthene	1.099	1.220	-11.0	101	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	119	0.00
17	Pyrene	1.313	1.178	10.3	108	0.00
18	Benzo(a)anthracene	0.943	1.147	-21.6#	148	0.00
19	Chrysene	1.222	1.108	9.3	109	0.00
20 I	Perylene-d12	1.000	1.000	0.0	144	0.00
21	Benzo(b)fluoranthene	1.092	1.153	-5.6	157#	0.00
22	Benzo(k)fluoranthene	1.192	1.095	8.1	142	0.00
23 C	Benzo(a)pyrene	0.923	1.085	-17.6	179#	0.00
24	Indeno(1,2,3-cd)pyrene	0.959	1.227	-27.9#	193#	0.00
25	Dibenzo(a,h)anthracene	0.761	1.031	-35.5#	201#	0.00
26	Benzo(g,h,i)perylene	0.930	1.031	-10.9	167#	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0