

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE070618\
 Data File : BE096852.D
 Acq On : 6 Jul 2018 13:36
 Operator : SJ/JU
 Sample : SST0.446
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SST0.446

Quant Time: Jul 06 14:58:48 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE070618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 06 14:58:27 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.40	152	2731	0.40	ng/ul	0.00
2) Naphthalene-d8	10.14	136	14078	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.03	164	7874	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.78	188	21214	0.40	ng/ul	0.00
16) Chrysene-d12	20.98	240	17094	0.40	ng/ul	0.00
20) Perylene-d12	23.21	264	14929	0.40	ng/ul	0.00

System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.74	152	9361	0.39	ng/ul	0.00
14) Fluoranthene-d10	18.81	212	20458	0.33	ng/ul	0.00

Target Compounds				Qvalue		
3) Naphthalene	10.19	128	13591	0.379	ng/ul#	92
5) 2-Methylnaphthalene	11.81	142	9445	0.405	ng/ul	98
7) Acenaphthylene	13.74	152	12811	0.375	ng/ul	96
8) Acenaphthene	14.09	153	10552	0.373	ng/ul#	94
9) Fluorene	15.08	166	10802	0.357	ng/ul#	92
11) Pentachlorophenol	16.43	266	464	0.563	ng/ul	99
12) Phenanthrene	16.82	178	21691	0.364	ng/ul#	89
13) Anthracene	16.91	178	18259	0.369	ng/ul#	92
15) Fluoranthene	18.84	202	22701	0.345	ng/ul	84
17) Pyrene	19.21	202	22250	0.372	ng/ul#	81
18) Benzo(a)anthracene	20.96	228	15771	0.385	ng/ul#	86
19) Chrysene	21.02	228	20595	0.359	ng/ul	97
21) Benzo(b)fluoranthene	22.54	252	15804	0.358	ng/ul	84
22) Benzo(k)fluoranthene	22.58	252	16633	0.363	ng/ul#	85
23) Benzo(a)pyrene	23.11	252	13063	0.330	ng/ul#	78
24) Indeno(1,2,3-cd)pyrene	25.50	276	13704	0.345	ng/ul#	80
25) Dibenzo(a,h)anthracene	25.53	278	11046	0.359	ng/ul#	88
26) Benzo(g,h,i)perylene	26.20	276	13272	0.337	ng/ul	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE070618\
Data File : BE096852.D
Acq On : 6 Jul 2018 13:36
Operator : SJ/JU
Sample : SST0.446
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTD0.446

Quant Time: Jul 06 14:58:48 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE070618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 06 14:58:27 2018
Response via : Initial Calibration

