

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE070618\
 Data File : BE096856.D
 Acq On : 6 Jul 2018 16:53
 Operator : SJ/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SICV50

Quant Time: Jul 06 17:30:38 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE070618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 06 16:55:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.40	152	3086	0.40	ng/ul	0.00
2) Naphthalene-d8	10.14	136	15176	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.03	164	8289	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.77	188	22100	0.40	ng/ul	0.00
16) Chrysene-d12	20.98	240	17581	0.40	ng/ul	0.00
20) Perylene-d12	23.22	264	14280	0.40	ng/ul	0.00

System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.74	152	9910	0.40	ng/ul	0.00
14) Fluoranthene-d10	18.81	212	20940	0.38	ng/ul	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Naphthalene	10.19	128	14764	0.402	ng/ul#	92
5) 2-Methylnaphthalene	11.81	142	10014	0.401	ng/ul	97
7) Acenaphthylene	13.74	152	13501	0.397	ng/ul	97
8) Acenaphthene	14.09	153	11103	0.396	ng/ul#	93
9) Fluorene	15.08	166	11482	0.397	ng/ul	93
11) Pentachlorophenol	16.43	266	475	0.328	ng/ul	90
12) Phenanthrene	16.82	178	22557	0.396	ng/ul#	88
13) Anthracene	16.91	178	18887	0.390	ng/ul#	92
15) Fluoranthene	18.85	202	23519	0.388	ng/ul	84
17) Pyrene	19.21	202	23901	0.414	ng/ul#	79
18) Benzo(a)anthracene	20.96	228	16352	0.394	ng/ul#	84
19) Chrysene	21.02	228	21491	0.400	ng/ul	97
21) Benzo(b)fluoranthene	22.54	252	16044	0.412	ng/ul	82
22) Benzo(k)fluoranthene	22.59	252	17062	0.401	ng/ul#	84
23) Benzo(a)pyrene	23.11	252	13160	0.399	ng/ul#	77
24) Indeno(1,2,3-cd)pyrene	25.51	276	13510	0.395	ng/ul#	79
25) Dibenzo(a,h)anthracene	25.54	278	10865	0.400	ng/ul#	85
26) Benzo(g,h,i)perylene	26.21	276	13440	0.405	ng/ul#	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE070618\
Data File : BE096856.D
Acq On : 6 Jul 2018 16:53
Operator : SJ/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SICV50

Quant Time: Jul 06 17:30:38 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE070618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 06 16:55:25 2018
Response via : Initial Calibration

