

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE070918\
 Data File : BE096877.D
 Acq On : 9 Jul 2018 12:16
 Operator : SJ/JU
 Sample : PB110794BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SBLK94

Quant Time: Jul 09 16:30:01 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE070618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jul 07 07:26:18 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.40	152	3049	0.40	ng/ul	0.00
2) Naphthalene-d8	10.14	136	16516	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.03	164	10697	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.88	188	2432331	0.40	ng/ul	0.10
16) Chrysene-d12	0.00	240	0	0.00	ng/ul	-20.98
20) Perylene-d12	23.22	264	43445	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.74	152	10666	0.39	ng/ul	0.00
14) Fluoranthene-d10	0.00	212	0	0.00	ng/ul	
Target Compounds						
9) Fluorene	15.04	166	3788	0.101	ng/ul#	17
23) Benzo(a)pyrene	23.08	252	9680	0.097	ng/ul#	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE070918\
Data File : BE096877.D
Acq On : 9 Jul 2018 12:16
Operator : SJ/JU
Sample : PB110794BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SBLK94

Quant Time: Jul 09 16:30:01 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE070618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jul 07 07:26:18 2018
Response via : Initial Calibration

