

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE071318\
 Data File : BE096900.D
 Acq On : 13 Jul 2018 11:05
 Operator : SJ/JU
 Sample : PB111075BS
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :

Quant Time: Jul 13 17:03:15 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 13 17:02:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.40	152	2056	0.40	ng	0.00
7) Naphthalene-d8	10.15	136	8695	0.40	ng	0.00
13) Acenaphthene-d10	14.03	164	3561	0.40	ng	0.00
19) Phenanthrene-d10	16.77	188	8208	0.40	ng	0.00
27) Chrysene-d12	20.98	240	8670	0.40	ng	0.00
34) Perylene-d12	23.21	264	9196	0.40	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.06	112	1997	0.43	ng	0.00
5) Phenol-d6	6.59	99	2515	0.37	ng	0.00
8) Nitrobenzene-d5	8.53	82	2486	0.47	ng	0.00
11) 2-Methylnaphthalene-d10	11.73	152	5551	0.44	ng	0.00
14) 2,4,6-Tribromophenol	15.52	330	215	0.34	ng	0.00
15) 2-Fluorobiphenyl	12.64	172	5894	0.45	ng	0.00
25) Fluoranthene-d10	18.81	212	32427	0.45	ng	0.00
29) Terphenyl-d14	19.44	244	8173	0.49	ng	0.00


Target Compounds				Qvalue		
2) 1,4-Dioxane	3.12	88	1099	0.357	ng	# 76
3) n-Nitrosodimethylamine	3.39	42	1410	0.432	ng	# 92
6) bis(2-Chloroethyl)ether	6.85	93	2702	0.359	ng	97
9) Naphthalene	10.20	128	32218	0.408	ng	100
10) Hexachlorobutadiene	10.50	225	1260	0.384	ng	98
12) 2-Methylnaphthalene	11.81	142	4947	0.378	ng	99
16) Acenaphthylene	13.74	152	6168	0.399	ng	99
17) Acenaphthene	14.08	154	3974	0.389	ng	94
18) Fluorene	15.08	166	4285	0.357	ng	# 81
20) 4-Bromophenyl-phenylether	15.99	248	1396	0.377	ng	# 73
21) Hexachlorobenzene	16.08	284	1598	0.379	ng	# 80
22) Pentachlorophenol	16.43	266	481	0.680	ng	# 73
23) Phenanthrene	16.81	178	8156	0.404	ng	98
24) Anthracene	16.91	178	7134	0.427	ng	97
26) Fluoranthene	18.84	202	9792	0.480	ng	98
28) Pyrene	19.21	202	9490	0.378	ng	99
30) Benzo(a)anthracene	20.95	228	8535	0.444	ng	98
31) Chrysene	21.01	228	9506	0.401	ng	99
32) Bis(2-ethylhexyl)phthalate	20.93	149	7091	0.412	ng	100
33) Indeno(1,2,3-cd)pyrene	25.51	276	8675	0.492	ng	# 81
35) Benzo(b)fluoranthene	22.54	252	8657	0.345	ng	# 88
36) Benzo(k)fluoranthene	22.58	252	9495	0.353	ng	95
37) Benzo(a)pyrene	23.11	252	7424	0.373	ng	92
38) Dibenzo(a,h)anthracene	25.54	278	7044	0.371	ng	97
39) Benzo(g,h,i)perylene	26.20	276	7816	0.351	ng	# 92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE071318\
Data File : BE096900.D
Acq On : 13 Jul 2018 11:05
Operator : SJ/JU
Sample : PB111075BS
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :

Quant Time: Jul 13 17:03:15 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 13 17:02:57 2018
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.40 min Scan# 394
Delta R.T. 0.01 min
Lab File: BE096900.D
Acq: 13 Jul 2018 11:05

Instrument :
BNA_E
ClientSampleId :

Tot Ion:152 Resp: 2056
Ion Ratio Lower Upper
152 100
150 162.6 117.2 175.8
115 69.5 57.0 85.6




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#2


1,4-Dioxane
Concen: 0.357 ng
RT: 3.12 min Scan# 22
Delta R.T. 0.02 min
Lab File: BE096900.D
Acq: 13 Jul 2018 11:05

Tgt Ion: 88 Resp: 1099
Ion Ratio Lower Upper
88 100
58 90.0 63.2 94.8
43 79.7 41.2 61.8#



Abundance Ion 88.00 (87.70 to 88.70): BE0
Ion 58.00 (57.70 to 58.70): BE0
Ion 43.00 (42.70 to 43.70): BE0





#3

n-Nitrosodimethylamine

Concen: 0.432 ng

RT: 3.39 min Scan# 46

Delta R.T. 0.01 min

Lab File: BE096900.D

Acq: 13 Jul 2018 11:05

Instrument :

BNA_E

ClientSampleId :


Tot Ion: 42 Resp: 1410

Ion Ratio Lower Upper




42 100

74 112.7 96.0 144.0

44 6.0 12.0 18.0#



Abundance Ion 42.00 (41.70 to 42.70): BE0
Ion 74.00 (73.70 to 74.70): BE0
Ion 44.00 (43.70 to 44.70): BE0



#4

2-Fluorophenol

Concen: 0.435 ng

RT: 5.06 min Scan# 191

Delta R.T. 0.01 min

Lab File: BE096900.D

Acq: 13 Jul 2018 11:05


Tgt Ion: 112 Resp: 1997

Ion Ratio Lower Upper



112 100


64 67.3 60.1 90.1

63 142.7 116.8 175.2



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BE0
Ion 63.00 (62.70 to 63.70): BE0





#5

Phenol-d6

Concen: 0.367 ng

RT: 6.59 min Scan# 324

Delta R.T. 0.01 min

Lab File: BE096900.D

Acq: 13 Jul 2018 11:05

Instrument :

BNA_E

ClientSampleId :


Tot Ion: 99 Resp: 2515

Ion Ratio Lower Upper

99 100

42 24.7 20.2 30.2




71 35.9 28.4 42.6



Abundance Ion 99.00 (98.70 to 99.70): BE0

Ion 42.00 (41.70 to 42.70): BE0

Ion 71.00 (70.70 to 71.70): BE0



#6

bis(2-Chloroethyl)ether

Concen: 0.359 ng

RT: 6.85 min Scan# 346

Delta R.T. 0.01 min

Lab File: BE096900.D

Acq: 13 Jul 2018 11:05


Tgt Ion: 93 Resp: 2702

Ion Ratio Lower Upper

93 100

63 338.3 275.3 412.9



95 32.8 25.2 37.8




Abundance Ion 93.00 (92.70 to 93.70): BE0

Ion 63.00 (62.70 to 63.70): BE0

Ion 95.00 (94.70 to 95.70): BE0





#7

Naphthalene-d8

Concen: 0.400 ng

RT: 10.15 min Scan# 613

Delta R.T. 0.00 min

Lab File: BE096900.D

Acq: 13 Jul 2018 11:05

Instrument :

BNA_E

ClientSampleId :

Tot Ion:136 Resp: 8695


Ion Ratio Lower Upper

136 100

137 11.4 9.5 14.3

54 48.4 29.1 43.7#

68 9.0 6.2 9.2




Abundance Ion 136.00 (135.70 to 136.70): E


Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BE0

Ion 68.00 (67.70 to 68.70): BE0



Time--> 10.00 10.10 10.20 10.30



#8

Nitrobenzene-d5

Concen: 0.472 ng

RT: 8.53 min Scan# 492

Delta R.T. 0.01 min

Lab File: BE096900.D

Acq: 13 Jul 2018 11:05


Tgt Ion: 82 Resp: 2486

Ion Ratio Lower Upper

82 100

128 33.7 24.0 36.0


54 51.9 40.0 60.0




Abundance Ion 82.00 (81.70 to 82.70): BE0

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BE0



Time--> 8.50 8.60



#9

Naphthalene

Concen: 0.408 ng

RT: 10.20 min Scan# 617

Delta R.T. 0.00 min

Lab File: BE096900.D

Acq: 13 Jul 2018 11:05

Instrument :

BNA_E

ClientSampleId :


Tot Ion:128 Resp: 32218

Ion Ratio Lower Upper

128 100

129 3.0 2.6 3.8




127 3.4 2.8 4.2



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#10

Hexachlorobutadiene

Concen: 0.384 ng

RT: 10.50 min Scan# 640

Delta R.T. 0.00 min

Lab File: BE096900.D

Acq: 13 Jul 2018 11:05


Tgt Ion:225 Resp: 1260

Ion Ratio Lower Upper

225 100

223 63.5 53.0 79.6



227 64.1 51.4 77.2




Abundance Ion 225.00 (224.70 to 225.70): E

Ion 223.00 (222.70 to 223.70): E

Ion 227.00 (226.70 to 227.70): E





#11

2-Methylnaphthalene-d10

Concen: 0.444 ng

RT: 11.73 min Scan# 762

Delta R.T. 0.00 min

Lab File: BE096900.D

Acq: 13 Jul 2018 11:05

Instrument :

BNA_E


ClientSampleId :

Tot Ion:152 Resp: 5551

Ion Ratio Lower Upper

152 100

151 15.6 15.4 23.0



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

11.73



3000

2000

1000

0

Time--> 11.60 11.70 11.80



#12

2-Methylnaphthalene

Concen: 0.378 ng

RT: 11.81 min Scan# 778

Delta R.T. 0.00 min

Lab File: BE096900.D

Acq: 13 Jul 2018 11:05


Tgt Ion:142 Resp: 4947

Ion Ratio Lower Upper

142 100

141 86.9 70.4 105.6

115 39.1 31.7 47.5



Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E

11.81

4000


3000


2000

1000

0

Time--> 11.80 11.90





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.03 min Scan# 1028

Delta R.T. -0.00 min

Lab File: BE096900.D

Acq: 13 Jul 2018 11:05

Instrument :

BNA_E

ClientSampleId :


Tot Ion:164 Resp: 3561

Ion Ratio Lower Upper

164 100

162 103.8 83.1 124.7




160 51.6 37.1 55.7



Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#14

2,4,6-Tribromophenol

Concen: 0.337 ng

RT: 15.52 min Scan# 1153

Delta R.T. 0.00 min

Lab File: BE096900.D

Acq: 13 Jul 2018 11:05


Tgt Ion:330 Resp: 215

Ion Ratio Lower Upper

330 100

332 102.3 152.7 229.1#



141 39.1 33.7 50.5




Abundance Ion 330.00 (329.70 to 330.70): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E









#15
2-Fluorobiphenyl
Concen: 0.452 ng
RT: 12.64 min Scan# 908
Delta R.T. -0.00 min
Lab File: BE096900.D
Acq: 13 Jul 2018 11:05

Instrument :
BNA_E
ClientSampleId :

Tot Ion:172 Resp: 5894
Ion Ratio Lower Upper
172 100
171 35.1 29.9 44.9
170 22.7 21.8 32.8




Abundance Ion 172.00 (171.70 to 172.70): E
Ion 171.00 (170.70 to 171.70): E
Ion 170.00 (169.70 to 170.70): E






#16
Acenaphthylene
Concen: 0.399 ng
RT: 13.74 min Scan# 1003
Delta R.T. -0.00 min
Lab File: BE096900.D
Acq: 13 Jul 2018 11:05

Tgt Ion:152 Resp: 6168
Ion Ratio Lower Upper
152 100
151 19.5 15.7 23.5
153 12.7 10.5 15.7



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 151.00 (150.70 to 151.70): E
Ion 153.00 (152.70 to 153.70): E





#17

Acenaphthene

Concen: 0.389 ng

RT: 14.08 min Scan# 1033

Delta R.T. -0.00 min

Lab File: BE096900.D

Acq: 13 Jul 2018 11:05

Instrument :

BNA_E

ClientSampleId :


Tot Ion:154 Resp: 3974

Ion Ratio Lower Upper

154 100

153 114.2 89.2 133.8




152 62.3 42.0 63.0



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#18

Fluorene

Concen: 0.357 ng

RT: 15.08 min Scan# 1115

Delta R.T. 0.00 min

Lab File: BE096900.D

Acq: 13 Jul 2018 11:05


Tgt Ion:166 Resp: 4285

Ion Ratio Lower Upper

166 100

165 97.6 77.8 116.6



167 14.6 42.4 63.6#




Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 16.77 min Scan# 1260
Delta R.T. 0.00 min
Lab File: BE096900.D
Acq: 13 Jul 2018 11:05




Instrument :
BNA_E
ClientSampleId :


Tot Ion:188 Resp: 8208
Ion Ratio Lower Upper
188 100
94 0.0 3.9 5.9#
80 15.2 3.6 5.4#



#20
4-Bromophenyl-phenylether
Concen: 0.377 ng
RT: 15.99 min Scan# 1193
Delta R.T. 0.00 min
Lab File: BE096900.D
Acq: 13 Jul 2018 11:05

Tgt Ion:248 Resp: 1396
Ion Ratio Lower Upper
248 100
250 107.6 62.7 94.1#
141 46.6 48.2 72.2#





#21

Hexachlorobenzene

Concen: 0.379 ng

RT: 16.08 min Scan# 1201

Delta R.T. -0.01 min

Lab File: BE096900.D

Acq: 13 Jul 2018 11:05

Instrument :

BNA_E

ClientSampleId :


Tot Ion:284 Resp: 1598

Ion Ratio Lower Upper

284 100

142 39.4 22.1 33.1#




249 31.5 34.8 52.2#



Abundance Ion 283.80 (283.50 to 284.50): E

Ion 142.00 (141.70 to 142.70): E

Ion 249.00 (248.70 to 249.70): E



#22

Pentachlorophenol

Concen: 0.680 ng

RT: 16.43 min Scan# 1231

Delta R.T. 0.00 min

Lab File: BE096900.D

Acq: 13 Jul 2018 11:05


Tgt Ion:266 Resp: 481

Ion Ratio Lower Upper

266 100

264 67.4 72.5 108.7#



268 64.9 73.8 110.6#




Abundance Ion 266.00 (265.70 to 266.70): E

Ion 264.00 (263.70 to 264.70): E

Ion 268.00 (267.70 to 268.70): E





#23

Phenanthrene

Concen: 0.404 ng

RT: 16.81 min Scan# 1264

Delta R.T. 0.00 min

Lab File: BE096900.D

Acq: 13 Jul 2018 11:05

Instrument :

BNA_E

ClientSampleId :





Tot Ion:178 Resp: 8156

Ion Ratio Lower Upper

178 100

176 19.7 15.1 22.7

179 15.4 11.8 17.6



#24

Anthracene

Concen: 0.427 ng

RT: 16.91 min Scan# 1272

Delta R.T. 0.00 min

Lab File: BE096900.D

Acq: 13 Jul 2018 11:05




Tgt Ion:178 Resp: 7134


Ion Ratio Lower Upper

178 100

176 18.3 12.8 19.2

179 15.8 13.0 19.6





#25

Fluoranthene-d10

Concen: 0.449 ng

RT: 18.81 min Scan# 1502

Delta R.T. 0.00 min

Lab File: BE096900.D

Acq: 13 Jul 2018 11:05

Instrument :

BNA_E

ClientSampleId :


Tot Ion:212 Resp: 32427

Ion Ratio Lower Upper

212 100

106 3.5 2.8 4.2




104 1.9 1.6 2.4



Abundance Ion 212.00 (211.70 to 212.70): E

Ion 106.00 (105.70 to 106.70): E

Ion 104.00 (103.70 to 104.70): E



#26

Fluoranthene

Concen: 0.480 ng

RT: 18.84 min Scan# 1507

Delta R.T. 0.00 min

Lab File: BE096900.D

Acq: 13 Jul 2018 11:05


Tgt Ion:202 Resp: 9792

Ion Ratio Lower Upper

202 100

101 10.8 9.8 14.8



203 16.9 13.7 20.5




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E





#27

Chrysene-d12

Concen: 0.400 ng

RT: 20.98 min Scan# 1786

Delta R.T. 0.00 min

Lab File: BE096900.D

Acq: 13 Jul 2018 11:05

Instrument :

BNA_E

ClientSampleId :


Tot Ion:240 Resp: 8670

Ion Ratio Lower Upper

240 100

120 6.4 5.5 8.3




236 27.3 20.7 31.1



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#28

Pyrene

Concen: 0.378 ng

RT: 19.21 min Scan# 1570

Delta R.T. -0.00 min

Lab File: BE096900.D

Acq: 13 Jul 2018 11:05


Tgt Ion:202 Resp: 9490

Ion Ratio Lower Upper

202 100

200 20.1 16.6 25.0



203 17.6 13.8 20.8




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E









#29
 Terphenyl-d14
 Concen: 0.488 ng
 RT: 19.44 min Scan# 1610
 Delta R.T. 0.00 min
 Lab File: BE096900.D
 Acq: 13 Jul 2018 11:05




Instrument :
 BNA_E
 ClientSampleId :


Tot Ion:244 Resp: 8173
 Ion Ratio Lower Upper
 244 100
 212 33.3 25.4 38.0
 122 10.7 8.1 12.1



#30
 Benzo(a)anthracene
 Concen: 0.444 ng
 RT: 20.95 min Scan# 1784
 Delta R.T. -0.01 min
 Lab File: BE096900.D
 Acq: 13 Jul 2018 11:05

Tgt Ion:228 Resp: 8535
 Ion Ratio Lower Upper
 228 100
 226 28.2 24.0 36.0
 229 19.8 16.0 24.0





#31

Chrysene

Concen: 0.401 ng

RT: 21.01 min Scan# 1789

Delta R.T. 0.00 min

Lab File: BE096900.D

Acq: 13 Jul 2018 11:05

Instrument :

BNA_E

ClientSampleId :


Tot Ion:228 Resp: 9506

Ion	Ratio	Lower	Upper
228	100		
226	29.2	24.0	36.0
229	20.2	16.0	24.0

228 100

226 29.2 24.0 36.0




229 20.2 16.0 24.0



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#32

Bis(2-ethylhexyl)phthalate

Concen: 0.412 ng

RT: 20.93 min Scan# 1782

Delta R.T. 0.00 min

Lab File: BE096900.D

Acq: 13 Jul 2018 11:05


Tgt Ion:149 Resp: 7091

Ion	Ratio	Lower	Upper
149	100		
167	6.8	5.4	8.0
279	0.9	0.7	1.1

149 100

167 6.8 5.4 8.0



279 0.9 0.7 1.1




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E









#33
 Indeno(1,2,3-cd)pyrene
 Concen: 0.492 ng
 RT: 25.51 min Scan# 2844
 Delta R.T. -0.00 min
 Lab File: BE096900.D
 Acq: 13 Jul 2018 11:05

Instrument :
 BNA_E
 ClientSampleId :

Tot Ion:276 Resp: 8675
 Ion Ratio Lower Upper
 276 100
 138 11.2 22.5 33.7#
 277 22.3 19.9 29.9




Abundance Ion 276.00 (275.70 to 276.70): E
 Ion 138.00 (137.70 to 138.70): E
 Ion 277.00 (276.70 to 277.70): E






#34
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.21 min Scan# 2204
 Delta R.T. 0.00 min
 Lab File: BE096900.D
 Acq: 13 Jul 2018 11:05

Tgt Ion:264 Resp: 9196
 Ion Ratio Lower Upper
 264 100
 260 26.6 20.5 30.7
 265 26.5 22.8 34.2



Abundance Ion 264.00 (263.70 to 264.70): E
 Ion 260.00 (259.70 to 260.70): E
 Ion 265.00 (264.70 to 265.70): E









#35
Benzo(b)fluoranthene
Concen: 0.345 ng
RT: 22.54 min Scan# 2014
Delta R.T. -0.00 min
Lab File: BE096900.D
Acq: 13 Jul 2018 11:05




Instrument :
BNA_E
ClientSampleId :


Tot Ion:252 Resp: 8657
Ion Ratio Lower Upper
252 100
253 22.5 20.0 30.0
125 10.7 16.0 24.0#



#36
Benzo(k)fluoranthene
Concen: 0.353 ng
RT: 22.58 min Scan# 2027
Delta R.T. 0.00 min
Lab File: BE096900.D
Acq: 13 Jul 2018 11:05

Tgt Ion:252 Resp: 9495
Ion Ratio Lower Upper
252 100
253 22.9 16.0 24.0
125 10.7 8.0 12.0









#37
Benzo(a)pyrene
Concen: 0.373 ng
RT: 23.11 min Scan# 2176
Delta R.T. 0.00 min
Lab File: BE096900.D
Acq: 13 Jul 2018 11:05

Instrument :
BNA_E
ClientSampleId :

Tot Ion:252 Resp: 7424
Ion Ratio Lower Upper
252 100
253 24.0 16.0 24.0
125 12.1 12.0 18.0




Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E






#38
Dibenzo(a,h)anthracene
Concen: 0.371 ng
RT: 25.54 min Scan# 2851
Delta R.T. 0.01 min
Lab File: BE096900.D
Acq: 13 Jul 2018 11:05

Tgt Ion:278 Resp: 7044
Ion Ratio Lower Upper
278 100
139 18.1 16.0 24.0
279 26.3 20.0 30.0



Abundance Ion 278.00 (277.70 to 278.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 279.00 (278.70 to 279.70): E





#39

Benzo(a,h,i)perylene

Concen: 0.351 ng

RT: 26.20 min Scan# 3038

Delta R.T. -0.00 min


Lab File: BE096900.D

Acq: 13 Jul 2018 11:05

Instrument :

BNA_E

ClientSampleId :



Tot Ion: 276 Resp: 7816

Ion Ratio Lower Upper

276 100

277 24.7 16.0 24.0#

138 22.2 20.0 30.0

Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

