

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE071318\
 Data File : BE096910.D
 Acq On : 13 Jul 2018 17:16
 Operator : SJ/JU
 Sample : J3170-08
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 RE120D2-071118

Quant Time: Jul 13 18:02:00 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 13 17:02:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.40	152	2042	0.40	ng	0.01
7) Naphthalene-d8	10.15	136	10230	0.40	ng	0.00
13) Acenaphthene-d10	14.02	164	5808	0.40	ng	0.00
19) Phenanthrene-d10	16.77	188	15814	0.40	ng	0.00
27) Chrysene-d12	20.98	240	14982	0.40	ng	0.00
34) Perylene-d12	23.21	264	15051	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.06	112	1141	0.25	ng	0.00
5) Phenol-d6	6.59	99	945	0.14	ng	0.00
8) Nitrobenzene-d5	8.52	82	3511	0.57	ng	0.00
11) 2-Methylnaphthalene-d10	11.73	152	6817	0.46	ng	0.00
14) 2,4,6-Tribromophenol	15.52	330	820	0.57	ng	0.00
15) 2-Fluorobiphenyl	12.64	172	12059	0.57	ng	0.00
25) Fluoranthene-d10	18.81	212	76946	0.55	ng	0.00
29) Terphenyl-d14	19.43	244	20103	0.70	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.10	88	18391	6.010	ng	Qvalue # 90
6) bis(2-Chloroethyl)ether	6.84	93	214	0.029	ng	85
32) Bis(2-ethylhexyl)phthalate	20.93	149	36757	1.236	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE071318\
Data File : BE096910.D
Acq On : 13 Jul 2018 17:16
Operator : SJ/JU
Sample : J3170-08
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
RE120D2-071118

Quant Time: Jul 13 18:02:00 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 13 17:02:57 2018
Response via : Initial Calibration

