

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE071318\
 Data File : BE096923.D
 Acq On : 14 Jul 2018 3:47
 Operator : SJ/JU
 Sample : J3936-03
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 OUTFALL001-180619

Quant Time: Jul 14 04:42:19 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 13 17:02:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.40	152	2029	0.40	ng	0.00
7) Naphthalene-d8	10.15	136	10200	0.40	ng	0.00
13) Acenaphthene-d10	14.03	164	5992	0.40	ng	0.00
19) Phenanthrene-d10	16.77	188	15212	0.40	ng	0.00
27) Chrysene-d12	20.98	240	13131	0.40	ng	0.00
34) Perylene-d12	23.21	264	13381	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.05	112	880	0.19	ng	0.00
5) Phenol-d6	6.59	99	734	0.11	ng	0.00
8) Nitrobenzene-d5	8.53	82	3051	0.49	ng	0.00
11) 2-Methylnaphthalene-d10	11.73	152	6147	0.42	ng	0.00
14) 2,4,6-Tribromophenol	15.52	330	721	0.51	ng	0.00
15) 2-Fluorobiphenyl	12.64	172	8930	0.41	ng	0.00
25) Fluoranthene-d10	18.81	212	59080	0.44	ng	0.00
29) Terphenyl-d14	19.43	244	17346	0.68	ng	0.00
Target Compounds						
9) Naphthalene	10.20	128	3888	0.042	ng	Qvalue # 93
32) Bis(2-ethylhexyl)phthalate	20.93	149	2799	0.107	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE071318\
Data File : BE096923.D
Acq On : 14 Jul 2018 3:47
Operator : SJ/JU
Sample : J3936-03
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
OUTFALL001-180619

Quant Time: Jul 14 04:42:19 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 13 17:02:57 2018
Response via : Initial Calibration

