

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE071318\
 Data File : BE096899.D
 Acq On : 13 Jul 2018 10:26
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 13 17:03:02 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 13 17:02:57 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
2	1,4-Dioxane	0.599	0.505	15.7	70	0.00
3	n-Nitrosodimethylamine	0.635	0.633	0.3	90	0.00
4 S	2-Fluorophenol	0.894	0.930	-4.0	97	0.00
5 S	Phenol-d6	1.334	1.418	-6.3	100	0.00
6	bis(2-Chloroethyl)ether	1.462	1.388	5.1	85	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	96	0.00
8 S	Nitrobenzene-d5	0.242	0.275	-13.6	118	0.00
9	Naphthalene	3.630	3.580	1.4	96	0.00
10	Hexachlorobutadiene	0.151	0.141	6.6	91	0.00
11 SURR	2-Methylnaphthalene-d10	0.576	0.587	-1.9	99	0.00
12	2-Methylnaphthalene	0.602	0.600	0.3	97	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	94	0.00
14 S	2,4,6-Tribromophenol	0.095	0.086	9.5	100	0.00
15 S	2-Fluorobiphenyl	1.464	1.483	-1.3	94	0.00
16	Acenaphthylene	1.734	1.758	-1.4	98	0.00
17	Acenaphthene	1.147	1.124	2.0	91	0.00
18	Fluorene	1.348	1.244	7.7	89	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	82	0.00
20	4-Bromophenyl-phenylether	0.180	0.185	-2.8	85	0.00
21	Hexachlorobenzene	0.206	0.216	-4.9	84	0.00
22	Pentachlorophenol	0.038	0.037	2.6	102	0.00
23	Phenanthrene	0.983	0.958	2.5	81	0.00
24	Anthracene	0.815	0.815	0.0	87	0.00
25 SURR	Fluoranthene-d10	3.521	3.494	0.8	85	0.00
26	Fluoranthene	0.994	1.010	-1.6	82	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	87	0.00
28	Pvrene	1.157	1.159	-0.2	89	0.00
29 S	Terphenyl-d14	0.772	0.751	2.7	86	0.00
30	Benzo(a)anthracene	0.886	0.896	-1.1	96	0.00
31	Chrysene	1.095	1.027	6.2	83	0.00
32	Bis(2-ethylhexyl)phthalate	0.794	0.804	-1.3	99	0.00
33	Indeno(1,2,3-cd)pyrene	0.814	0.793	2.6	92	0.00
34 I	Perylene-d12	1.000	1.000	0.0	99	0.00
35	Benzo(b)fluoranthene	1.092	0.968	11.4	93	0.00
36	Benzo(k)fluoranthene	1.233	1.112	9.8	96	0.00
37 C	Benzo(a)pyrene	0.867	0.858	1.0	104	0.00
38	Dibenzo(a,h)anthracene	0.889	0.745	16.2	91	0.00
39	Benzo(a,h,i)perylene	0.969	0.814	16.0	90	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE071318\
Data File : BE096899.D
Acq On : 13 Jul 2018 10:26
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 13 17:03:02 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 13 17:02:57 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			