

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE071318\  
 Data File : BE096916.D  
 Acq On : 13 Jul 2018 20:58  
 Operator : SJ/JU  
 Sample : SSTDCCC0.4  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_E  
 LabSampleId :  
 SSTDCCC0.4

Quant Time: Jul 13 22:58:59 2018  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062018.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Jul 13 17:02:57 2018  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
2	1,4-Dioxane	0.599	0.532	11.2	77	0.00
3	n-Nitrosodimethylamine	0.635	0.649	-2.2	96	0.00
4 S	2-Fluorophenol	0.894	1.020	-14.1	111	0.00
5 S	Phenol-d6	1.334	1.576	-18.1	116	0.00
6	bis(2-Chloroethyl)ether	1.462	1.415	3.2	90	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	99	0.01
8 S	Nitrobenzene-d5	0.242	0.293	-21.1	129	0.00
9	Naphthalene	3.630	3.532	2.7	98	0.00
10	Hexachlorobutadiene	0.151	0.146	3.3	97	0.00
11 SURR	2-Methylnaphthalene-d10	0.576	0.577	-0.2	100	0.00
12	2-Methylnaphthalene	0.602	0.596	1.0	99	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	96	0.00
14 S	2,4,6-Tribromophenol	0.095	0.096	-1.1	114	0.00
15 S	2-Fluorobiphenyl	1.464	1.441	1.6	93	0.00
16	Acenaphthylene	1.734	1.770	-2.1	101	0.00
17	Acenaphthene	1.147	1.121	2.3	93	0.00
18	Fluorene	1.348	1.259	6.6	92	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	84	0.01
20	4-Bromophenyl-phenylether	0.180	0.185	-2.8	87	0.00
21	Hexachlorobenzene	0.206	0.207	-0.5	82	0.00
22	Pentachlorophenol	0.038	0.029	23.7	83	0.00
23	Phenanthrene	0.983	0.963	2.0	83	0.00
24	Anthracene	0.815	0.843	-3.4	91	0.00
25 SURR	Fluoranthene-d10	3.521	3.588	-1.9	89	0.00
26	Fluoranthene	0.994	1.033	-3.9	86	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	91	0.00
28	Pyrene	1.157	1.137	1.7	91	0.00
29 S	Terphenyl-d14	0.772	0.730	5.4	88	0.00
30	Benzo(a)anthracene	0.886	0.937	-5.8	105	0.00
31	Chrysene	1.095	1.021	6.8	86	0.00
32	Bis(2-ethylhexyl)phthalate	0.794	1.178	-48.4#	151#	0.00
33	Indeno(1,2,3-cd)pyrene	0.814	0.991	-21.7	120	0.00
34 I	Perylene-d12	1.000	1.000	0.0	123	0.00
35	Benzo(b)fluoranthene	1.092	0.879	19.5	104	0.00
36	Benzo(k)fluoranthene	1.233	0.989	19.8	105	0.00
37 C	Benzo(a)pyrene	0.867	0.806	7.0	121	0.00
38	Dibenzo(a,h)anthracene	0.889	0.789	11.2	120	0.00
39	Benzo(a,h,i)perylene	0.969	0.818	15.6	112	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE071318\  
Data File : BE096916.D  
Acq On : 13 Jul 2018 20:58  
Operator : SJ/JU  
Sample : SSTDCCC0.4  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
BNA\_E  
LabSampleId :  
SSTDCCC0.4

Quant Time: Jul 13 22:58:59 2018  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062018.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Jul 13 17:02:57 2018  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
-----				
(#) = Out of Range	SPCC's out = 0	CCC's out = 0		