

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE071318\
 Data File : BE096918.D
 Acq On : 14 Jul 2018 00:42
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 16 01:03:34 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 13 17:02:57 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	90	0.00
2	1,4-Dioxane	0.400	0.390	2.5	84	0.00
3	n-Nitrosodimethylamine	0.400	0.414	-3.5	96	0.00
4 S	2-Fluorophenol	0.400	0.492	-23.0	119	0.00
5 S	Phenol-d6	0.400	0.501	-25.2#	122	0.00
6	bis(2-Chloroethyl)ether	0.400	0.401	-0.3	93	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	102	0.00
8 S	Nitrobenzene-d5	0.400	0.486	-21.5	133	0.00
9	Naphthalene	0.400	0.387	3.3	99	0.00
10	Hexachlorobutadiene	0.400	0.371	7.3	95	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.410	-2.5	104	0.00
12	2-Methylnaphthalene	0.400	0.405	-1.3	104	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	100	0.00
14 S	2,4,6-Tribromophenol	0.400	0.490	-22.5	141	0.00
15 S	2-Fluorobiphenyl	0.400	0.399	0.3	98	0.00
16	Acenaphthylene	0.400	0.417	-4.2	107	0.00
17	Acenaphthene	0.400	0.399	0.3	98	0.00
18	Fluorene	0.400	0.380	5.0	97	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	88	0.01
20	4-Bromophenyl-phenylether	0.400	0.415	-3.7	93	0.00
21	Hexachlorobenzene	0.400	0.406	-1.5	87	0.00
22	Pentachlorophenol	0.400	0.581	-45.2#	144	0.00
23	Phenanthrene	0.400	0.392	2.0	87	0.00
24	Anthracene	0.400	0.410	-2.5	95	0.00
25 SURR	Fluoranthene-d10	0.400	0.406	-1.5	94	0.00
26	Fluoranthene	0.400	0.413	-3.2	90	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	98	0.00
28	Pvrene	0.400	0.402	-0.5	100	0.00
29 S	Terphenyl-d14	0.400	0.382	4.5	95	0.00
30	Benzo(a)anthracene	0.400	0.431	-7.7	116	-0.01
31	Chrysene	0.400	0.366	8.5	91	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.636	-59.0#	174	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.497	-24.2	132	0.00
34 I	Perylene-d12	0.400	0.400	0.0	133	0.00
35	Benzo(b)fluoranthene	0.400	0.318	20.5	111	0.00
36	Benzo(k)fluoranthene	0.400	0.345	13.8	115	0.00
37 C	Benzo(a)pyrene	0.400	0.374	6.5	132	0.00
38	Dibenzo(a,h)anthracene	0.400	0.374	6.5	127	0.00
39	Benzo(a,h,i)perylene	0.400	0.340	15.0	121	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE071318\
Data File : BE096918.D
Acq On : 14 Jul 2018 00:42
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 16 01:03:34 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 13 17:02:57 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				