

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE071618\
 Data File : BE096941.D
 Acq On : 16 Jul 2018 16:24
 Operator : SJ/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE071618

Quant Time: Jul 17 15:09:42 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE071618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 17 15:02:56 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2	1,4-Dioxane	0.573	0.493	14.0	89	0.00
3	n-Nitrosodimethylamine	0.644	0.559	13.2	92	0.00
4 S	2-Fluorophenol	0.957	0.800	16.4	94	0.00
5 S	Phenol-d6	1.460	1.236	15.3	92	0.00
6	bis(2-Chloroethyl)ether	1.433	1.254	12.5	90	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	98	0.00
8 S	Nitrobenzene-d5	0.292	0.252	13.7	91	0.00
9	Naphthalene	3.581	3.169	11.5	91	0.00
10	Hexachlorobutadiene	0.145	0.128	11.7	91	0.00
11 SURR	2-Methylnaphthalene-d10	0.583	0.503	13.7	89	0.00
12	2-Methylnaphthalene	0.607	0.524	13.7	89	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	97	0.00
14 S	2,4,6-Tribromophenol	0.101	0.076	24.8	89	0.00
15 S	2-Fluorobiphenyl	1.488	1.347	9.5	90	0.00
16	Acenaphthylene	1.827	1.564	14.4	87	0.00
17	Acenaphthene	1.136	0.999	12.1	89	0.00
18	Fluorene	1.275	1.105	13.3	89	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	100	-0.01
20	4-Bromophenyl-phenylether	0.189	0.162	14.3	90	-0.01
21	Hexachlorobenzene	0.210	0.184	12.4	90	-0.01
22	Pentachlorophenol	0.034	0.027	20.6	88	0.00
23	Phenanthrene	0.958	0.829	13.5	90	0.00
24	Anthracene	0.834	0.706	15.3	91	0.00
25 SURR	Fluoranthene-d10	3.480	2.854	18.0	89	0.00
26	Fluoranthene	1.010	0.831	17.7	88	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	96	0.00
28	Pvrene	1.181	1.037	12.2	88	0.00
29 S	Terphenyl-d14	0.772	0.684	11.4	90	0.00
30	Benzo(a)anthracene	0.952	0.798	16.2	88	-0.01
31	Chrysene	1.068	0.936	12.4	86	0.00
32	Bis(2-ethylhexyl)phthalate	0.951	0.749	21.2	87	0.00
33	Indeno(1,2,3-cd)pyrene	0.778	0.629	19.2	87	0.00
34 I	Perylene-d12	1.000	1.000	0.0	99	0.00
35	Benzo(b)fluoranthene	1.021	0.828	18.9	87	0.00
36	Benzo(k)fluoranthene	1.214	0.991	18.4	85	0.00
37 C	Benzo(a)pyrene	0.913	0.707	22.6#	85	0.00
38	Dibenzo(a,h)anthracene	0.746	0.596	20.1	90	0.00
39	Benzo(a,h,i)perylene	0.830	0.701	15.5	93	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE071618\
Data File : BE096941.D
Acq On : 16 Jul 2018 16:24
Operator : SJ/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE071618

Quant Time: Jul 17 15:09:42 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE071618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 17 15:02:56 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 1			