

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE071618\
 Data File : BE096943.D
 Acq On : 16 Jul 2018 18:09
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 17 15:14:38 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE071618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 17 15:02:56 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.573	0.553	3.5	100	0.00
3	n-Nitrosodimethylamine	0.644	0.649	-0.8	107	0.00
4 S	2-Fluorophenol	0.957	0.960	-0.3	114	0.00
5 S	Phenol-d6	1.460	1.525	-4.5	114	0.00
6	bis(2-Chloroethyl)ether	1.433	1.451	-1.3	105	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	99	0.00
8 S	Nitrobenzene-d5	0.292	0.307	-5.1	112	0.00
9	Naphthalene	3.581	3.603	-0.6	104	0.00
10	Hexachlorobutadiene	0.145	0.147	-1.4	105	0.00
11 SURR	2-Methylnaphthalene-d10	0.583	0.572	1.9	102	0.00
12	2-Methylnaphthalene	0.607	0.608	-0.2	104	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	98	0.00
14 S	2,4,6-Tribromophenol	0.101	0.100	1.0	118	0.00
15 S	2-Fluorobiphenyl	1.488	1.576	-5.9	107	0.00
16	Acenaphthylene	1.827	1.882	-3.0	105	0.00
17	Acenaphthene	1.136	1.161	-2.2	105	0.00
18	Fluorene	1.275	1.294	-1.5	105	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	99	-0.01
20	4-Bromophenyl-phenylether	0.189	0.191	-1.1	106	-0.01
21	Hexachlorobenzene	0.210	0.214	-1.9	104	-0.01
22	Pentachlorophenol	0.034	0.036	-5.9	118	0.00
23	Phenanthrene	0.958	0.973	-1.6	105	0.00
24	Anthracene	0.834	0.860	-3.1	110	0.00
25 SURR	Fluoranthene-d10	3.480	3.314	4.8	102	0.00
26	Fluoranthene	1.010	0.999	1.1	105	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	104	0.00
28	Pvrene	1.181	1.186	-0.4	109	0.00
29 S	Terphenyl-d14	0.772	0.797	-3.2	113	0.00
30	Benzo(a)anthracene	0.952	0.937	1.6	112	-0.01
31	Chrysene	1.068	1.051	1.6	105	0.00
32	Bis(2-ethylhexyl)phthalate	0.951	0.937	1.5	118	0.00
33	Indeno(1,2,3-cd)pyrene	0.778	0.823	-5.8	124	0.00
34 I	Perylene-d12	1.000	1.000	0.0	120	0.00
35	Benzo(b)fluoranthene	1.021	0.894	12.4	113	0.00
36	Benzo(k)fluoranthene	1.214	1.077	11.3	111	0.00
37 C	Benzo(a)pyrene	0.913	0.789	13.6	115	0.00
38	Dibenzo(a,h)anthracene	0.746	0.718	3.8	131	0.00
39	Benzo(a,h,i)perylene	0.830	0.769	7.3	123	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE071618\
Data File : BE096943.D
Acq On : 16 Jul 2018 18:09
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 17 15:14:38 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE071618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 17 15:02:56 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			