

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE071618\
 Data File : BE096950.D
 Acq On : 16 Jul 2018 22:41
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 17 15:33:01 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE071618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 17 15:02:56 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	1,4-Dioxane	0.573	0.543	5.2	99	0.00
3	n-Nitrosodimethylamine	0.644	0.644	0.0	108	0.00
4 S	2-Fluorophenol	0.957	1.028	-7.4	123	0.00
5 S	Phenol-d6	1.460	1.622	-11.1	123	0.00
6	bis(2-Chloroethyl)ether	1.433	1.437	-0.3	105	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
8 S	Nitrobenzene-d5	0.292	0.309	-5.8	114	0.00
9	Naphthalene	3.581	3.639	-1.6	107	0.00
10	Hexachlorobutadiene	0.145	0.143	1.4	104	0.00
11 SURR	2-Methylnaphthalene-d10	0.583	0.582	0.2	106	0.00
12	2-Methylnaphthalene	0.607	0.623	-2.6	108	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	102	0.00
14 S	2,4,6-Tribromophenol	0.101	0.113	-11.9	139	0.00
15 S	2-Fluorobiphenyl	1.488	1.580	-6.2	111	0.00
16	Acenaphthylene	1.827	1.873	-2.5	109	0.00
17	Acenaphthene	1.136	1.188	-4.6	112	0.00
18	Fluorene	1.275	1.301	-2.0	110	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	101	-0.01
20	4-Bromophenyl-phenylether	0.189	0.193	-2.1	110	0.00
21	Hexachlorobenzene	0.210	0.218	-3.8	109	-0.01
22	Pentachlorophenol	0.034	0.045	-32.4#	151#	-0.01
23	Phenanthrene	0.958	0.991	-3.4	110	0.00
24	Anthracene	0.834	0.855	-2.5	112	0.00
25 SURR	Fluoranthene-d10	3.480	3.486	-0.2	110	0.00
26	Fluoranthene	1.010	1.052	-4.2	113	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	114	0.00
28	Pvrene	1.181	1.189	-0.7	119	0.00
29 S	Terphenyl-d14	0.772	0.792	-2.6	123	0.00
30	Benzo(a)anthracene	0.952	0.980	-2.9	128	-0.01
31	Chrysene	1.068	1.045	2.2	114	0.00
32	Bis(2-ethylhexyl)phthalate	0.951	1.129	-18.7	155#	0.00
33	Indeno(1,2,3-cd)pyrene	0.778	0.892	-14.7	146	0.00
34 I	Perylene-d12	1.000	1.000	0.0	138	0.00
35	Benzo(b)fluoranthene	1.021	0.882	13.6	128	0.00
36	Benzo(k)fluoranthene	1.214	1.089	10.3	129	0.00
37 C	Benzo(a)pyrene	0.913	0.821	10.1	137	0.00
38	Dibenzo(a,h)anthracene	0.746	0.755	-1.2	158#	0.00
39	Benzo(a,h,i)perylene	0.830	0.778	6.3	143	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE071618\
Data File : BE096950.D
Acq On : 16 Jul 2018 22:41
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 17 15:33:01 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE071618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 17 15:02:56 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			