


Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE071618\
Data File : BE096942.D
Acq On : 16 Jul 2018 17:07
Operator : SJ/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE071618.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue Jul 17 15:02:56 2018



AutoFind: Scans 2495, 2496, 2497; Background Corrected with Scan 2488

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	65.9	19562	PASS
68	69	0.00	2	1.9	331	PASS
69	198	0.00	100	59.6	17688	PASS
70	69	0.00	2	0.5	85	PASS
127	198	10	80	63.1	18723	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	29676	PASS
199	198	5	9	6.6	1964	PASS
275	198	10	60	26.2	7779	PASS
365	198	1	100	3.9	1170	PASS
441	443	0.01	100	87.1	4462	PASS
442	198	50	100	89.4	26532	PASS
443	442	15	24	19.3	5125	PASS