

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE071718\
 Data File : BE096961.D
 Acq On : 17 Jul 2018 17:18
 Operator : SJ/JU
 Sample : J3967-03
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 DB1D7

Quant Time: Jul 18 10:50:57 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE071718.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 18 03:13:50 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.40	152	1609	0.40	ng/ul	0.00
2) Naphthalene-d8	10.14	136	7457	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.03	164	3977	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.78	188	11417	0.40	ng/ul	0.00
16) Chrysene-d12	20.98	240	12792	0.40	ng/ul	0.00
20) Perylene-d12	23.21	264	13061	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.74	152	4443	0.38	ng/ul	0.00
14) Fluoranthene-d10	18.81	212	13328	0.38	ng/ul	0.00
Target Compounds					Ovalue	
3) Naphthalene	10.19	128	7361	0.407	ng/ul#	93
5) 2-Methylnaphthalene	11.81	142	10443	0.875	ng/ul	98
7) Acenaphthylene	13.74	152	9782	0.564	ng/ul	95
8) Acenaphthene	14.09	153	17245	1.282	ng/ul#	94
11) Pentachlorophenol	16.43	266	910	0.705	ng/ul	99
12) Phenanthrene	16.81	178	16582	0.553	ng/ul#	89
15) Fluoranthene	18.84	202	34719	0.879	ng/ul	84
17) Pyrene	19.21	202	27524	0.744	ng/ul#	82
18) Benzo(a)anthracene	20.96	228	14840	0.438	ng/ul#	85
21) Benzo(b)fluoranthene	22.53	252	25443	0.686	ng/ul	83
22) Benzo(k)fluoranthene	22.58	252	31278	0.817	ng/ul#	84
23) Benzo(a)pyrene	23.11	252	13903	0.427	ng/ul#	78
24) Indeno(1,2,3-cd)pyrene	25.51	276	26045	0.714	ng/ul#	83
25) Dibenzo(a,h)anthracene	25.53	278	27531	0.904	ng/ul#	86
26) Benzo(g,h,i)perylene	26.20	276	29130	0.895	ng/ul#	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE071718\
Data File : BE096961.D
Acq On : 17 Jul 2018 17:18
Operator : SJ/JU
Sample : J3967-03
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
DB1D7

Quant Time: Jul 18 10:50:57 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE071718.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 18 03:13:50 2018
Response via : Initial Calibration

