

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE071718\
 Data File : BE096954.D
 Acq On : 17 Jul 2018 11:43
 Operator : SJ/JU
 Sample : SST00.262
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SST00.262

Quant Time: Jul 17 13:01:45 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE071718.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 17 13:01:30 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.39	152	1516	0.40	ng/ul	0.00
2) Naphthalene-d8	10.14	136	6847	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.03	164	3541	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.77	188	10308	0.40	ng/ul	0.00
16) Chrysene-d12	20.98	240	12297	0.40	ng/ul	0.00
20) Perylene-d12	23.21	264	11695	0.40	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	11.74	152	2152	0.19	ng/ul	0.00
14) Fluoranthene-d10	18.81	212	6119	0.24	ng/ul	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
3) Naphthalene	10.19	128	3336	0.202	ng/ul#	95
5) 2-Methylnaphthalene	11.81	142	2169	0.192	ng/ul	97
7) Acenaphthylene	13.74	152	3067	0.211	ng/ul	98
8) Acenaphthene	14.08	153	2361	0.197	ng/ul#	91
9) Fluorene	15.08	166	2496	0.202	ng/ul	96
11) Pentachlorophenol	16.43	266	213	0.316	ng/ul	93
12) Phenanthrene	16.82	178	5432	0.204	ng/ul#	90
13) Anthracene	16.90	178	4592	0.203	ng/ul#	91
15) Fluoranthene	18.84	202	7028	0.248	ng/ul	85
17) Pyrene	19.21	202	7539	0.187	ng/ul#	84
18) Benzo(a)anthracene	20.96	228	6435	0.222	ng/ul#	85
19) Chrysene	21.02	228	7559	0.201	ng/ul	98
21) Benzo(b)fluoranthene	22.54	252	6488	0.203	ng/ul	84
22) Benzo(k)fluoranthene	22.58	252	6481	0.186	ng/ul#	87
23) Benzo(a)pyrene	23.11	252	5625	0.208	ng/ul#	81
24) Indeno(1,2,3-cd)pyrene	25.50	276	6194	0.221	ng/ul#	80
25) Dibenzo(a,h)anthracene	25.53	278	5151	0.231	ng/ul	92
26) Benzo(g,h,i)perylene	26.20	276	5650	0.208	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE071718\
Data File : BE096954.D
Acq On : 17 Jul 2018 11:43
Operator : SJ/JU
Sample : SST0.262
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTD0.262

Quant Time: Jul 17 13:01:45 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE071718.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 17 13:01:30 2018
Response via : Initial Calibration

