

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE071818\
 Data File : BE096997.D
 Acq On : 18 Jul 2018 17:16
 Operator : SJ/JU
 Sample : J3967-18DL 5X
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 DB1F8DL

Manual Integrations
 APPROVED

Sohil
 7/19/2018 11:36:58 AM

Quant Time: Jul 18 17:57:54 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE071718.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 18 15:28:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.40	152	1491	0.40	ng/ul	0.00
2) Naphthalene-d8	10.14	136	7618	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.03	164	5145	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.78	188	13721	0.40	ng/ul	0.00
16) Chrysene-d12	20.98	240	13940	0.40	ng/ul	0.00
20) Perylene-d12	23.21	264	13281m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.74	152	795	0.07	ng/ul	0.00
14) Fluoranthene-d10	18.81	212	2949	0.07	ng/ul	0.00
Target Compounds						
9) Fluorene	15.08	166	379	0.020	ng/ul#	88
11) Pentachlorophenol	16.43	266	1461	0.942	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE071818\
Data File : BE096997.D
Acq On : 18 Jul 2018 17:16
Operator : SJ/JU
Sample : J3967-18DL 5X
Misc :
ALS Vial : 12 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
DB1F8DL

Quant Time: Jul 18 17:57:54 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE071718.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 18 15:28:33 2018
Response via : Initial Calibration

Manual Integrations
APPROVED

Sohil
7/19/2018 11:36:58 AM





#9

Fluorene

Concen: 0.020 ng/ul

RT: 15.08 min Scan# 2342

Delta R.T. -0.00 min

Lab File: BE096997.D

Acq: 18 Jul 2018 17:16

Instrument :

BNA_E





ClientSampled :

DB1F8DL

Tgt Ion	Ratio	Lower	Upper
166	100		
165	100.9	73.4	110.2
167	29.1	14.6	22.0

Manual Integrations
APPROVED

Sohil
7/19/2018 11:36:58 AM



#11

Pentachlorophenol

Concen: 0.942 ng/ul

RT: 16.43 min Scan# 2553

Delta R.T. 0.00 min

Lab File: BE096997.D

Acq: 18 Jul 2018 17:16

Tgt Ion	Ratio	Lower	Upper
266	100		
264	62.8	47.9	71.9
268	62.5	49.8	74.8

