

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE072018\
 Data File : BE097017.D
 Acq On : 20 Jul 2018 6:36
 Operator : SJ/JU
 Sample : J4014-13
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 GW03-071218

Quant Time: Jul 20 08:04:35 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE071618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 17 15:02:56 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.39	152	1531	0.40	ng	0.00
7) Naphthalene-d8	10.15	136	7029	0.40	ng	0.00
13) Acenaphthene-d10	14.01	164	4121	0.40	ng	-0.01
19) Phenanthrene-d10	16.77	188	12514	0.40	ng	-0.01
27) Chrysene-d12	20.98	240	12978	0.40	ng	0.00
34) Perylene-d12	23.21	264	12163	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.06	112	750	0.20	ng	0.00
5) Phenol-d6	6.60	99	634	0.11	ng	0.00
8) Nitrobenzene-d5	8.52	82	2207	0.43	ng	0.00
11) 2-Methylnaphthalene-d10	11.73	152	3893	0.38	ng	0.00
14) 2,4,6-Tribromophenol	15.52	330	486	0.52	ng	0.00
15) 2-Fluorobiphenyl	12.63	172	7030	0.46	ng	-0.01
25) Fluoranthene-d10	18.81	212	58940	0.54	ng	0.00
29) Terphenyl-d14	19.43	244	14243	0.57	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.10	88	1978	0.903	ng	91
6) bis(2-Chloroethyl)ether	6.84	93	196	0.036	ng	85
32) Bis(2-ethylhexyl)phthalate	20.93	149	5203	0.169	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE072018\
Data File : BE097017.D
Acq On : 20 Jul 2018 6:36
Operator : SJ/JU
Sample : J4014-13
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
GW03-071218

Quant Time: Jul 20 08:04:35 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE071618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 17 15:02:56 2018
Response via : Initial Calibration

