

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE072018\
 Data File : BE097016.D
 Acq On : 20 Jul 2018 5:57
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 20 06:39:06 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE071618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 17 15:02:56 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	81	0.00
2	1,4-Dioxane	0.573	0.583	-1.7	85	0.00
3	n-Nitrosodimethylamine	0.644	0.677	-5.1	90	0.00
4 S	2-Fluorophenol	0.957	1.035	-8.2	99	0.00
5 S	Phenol-d6	1.460	1.503	-2.9	91	0.00
6	bis(2-Chloroethyl)ether	1.433	1.419	1.0	83	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	74	0.00
8 S	Nitrobenzene-d5	0.292	0.309	-5.8	84	0.00
9	Naphthalene	3.581	3.443	3.9	74	0.00
10	Hexachlorobutadiene	0.145	0.157	-8.3	84	0.00
11 SURR	2-Methylnaphthalene-d10	0.583	0.553	5.1	74	0.00
12	2-Methylnaphthalene	0.607	0.578	4.8	73	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	88	0.00
14 S	2,4,6-Tribromophenol	0.101	0.113	-11.9	120	0.00
15 S	2-Fluorobiphenyl	1.488	1.306	12.2	79	0.00
16	Acenaphthylene	1.827	1.663	9.0	83	0.00
17	Acenaphthene	1.136	1.006	11.4	81	0.00
18	Fluorene	1.275	1.261	1.1	92	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	102	0.00
20	4-Bromophenyl-phenylether	0.189	0.181	4.2	104	0.00
21	Hexachlorobenzene	0.210	0.198	5.7	100	0.00
22	Pentachlorophenol	0.034	0.050	-47.1#	169#	0.00
23	Phenanthrene	0.958	0.919	4.1	103	0.00
24	Anthracene	0.834	0.789	5.4	104	0.00
25 SURR	Fluoranthene-d10	3.480	3.798	-9.1	121	0.00
26	Fluoranthene	1.010	1.060	-5.0	115	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	130	0.00
28	Pyrene	1.181	1.054	10.8	120	0.00
29 S	Terphenyl-d14	0.772	0.710	8.0	126	0.00
30	Benzo(a)anthracene	0.952	0.905	4.9	135	0.00
31	Chrysene	1.068	1.033	3.3	129	0.00
32	Bis(2-ethylhexyl)phthalate	0.951	0.870	8.5	137	0.00
33	Indeno(1,2,3-cd)pyrene	0.778	0.973	-25.1#	183#	0.00
34 I	Perylene-d12	1.000	1.000	0.0	150	0.00
35	Benzo(b)fluoranthene	1.021	0.836	18.1	132	0.00
36	Benzo(k)fluoranthene	1.214	1.017	16.2	131	0.00
37 C	Benzo(a)pyrene	0.913	0.793	13.1	144	0.00
38	Dibenzo(a,h)anthracene	0.746	0.885	-18.6	202#	0.00
39	Benzo(g,h,i)perylene	0.830	0.894	-7.7	179#	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE072018\
Data File : BE097016.D
Acq On : 20 Jul 2018 5:57
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 20 06:39:06 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE071618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 17 15:02:56 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		