

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE072018\
 Data File : BE097028.D
 Acq On : 20 Jul 2018 15:36
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 20 17:02:54 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE071618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 20 11:34:20 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	84	0.00
2	1,4-Dioxane	0.573	0.573	0.0	87	0.00
3	n-Nitrosodimethylamine	0.644	0.693	-7.6	96	0.00
4 S	2-Fluorophenol	0.957	0.975	-1.9	97	0.00
5 S	Phenol-d6	1.460	1.440	1.4	90	0.00
6	bis(2-Chloroethyl)ether	1.433	1.391	2.9	84	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	73	0.00
8 S	Nitrobenzene-d5	0.292	0.317	-8.6	86	0.00
9	Naphthalene	3.581	3.562	0.5	76	0.00
10	Hexachlorobutadiene	0.145	0.162	-11.7	86	0.01
11 SURR	2-Methylnaphthalene-d10	0.583	0.564	3.3	75	0.00
12	2-Methylnaphthalene	0.607	0.591	2.6	75	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	86	0.00
14 S	2,4,6-Tribromophenol	0.101	0.109	-7.9	113	0.00
15 S	2-Fluorobiphenyl	1.488	1.343	9.7	80	0.00
16	Acenaphthylene	1.827	1.680	8.0	83	0.00
17	Acenaphthene	1.136	1.028	9.5	82	0.00
18	Fluorene	1.275	1.323	-3.8	95	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
20	4-Bromophenyl-phenylether	0.189	0.186	1.6	104	0.00
21	Hexachlorobenzene	0.210	0.207	1.4	102	0.00
22	Pentachlorophenol	0.034	0.044	-29.4#	145	0.00
23	Phenanthrene	0.958	0.937	2.2	102	0.00
24	Anthracene	0.834	0.799	4.2	103	0.00
25 SURR	Fluoranthene-d10	3.480	3.814	-9.6	119	0.00
26	Fluoranthene	1.010	1.062	-5.1	112	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	124	0.00
28	Pyrene	1.181	1.028	13.0	112	0.00
29 S	Terphenyl-d14	0.772	0.718	7.0	122	0.00
30	Benzo(a)anthracene	0.952	0.905	4.9	129	0.00
31	Chrysene	1.068	1.057	1.0	126	0.00
32	Bis(2-ethylhexyl)phthalate	0.951	0.822	13.6	124	-0.01
33	Indeno(1,2,3-cd)pyrene	0.778	0.960	-23.4	173#	0.00
34 I	Perylene-d12	1.000	1.000	0.0	140	0.00
35	Benzo(b)fluoranthene	1.021	0.842	17.5	124	0.00
36	Benzo(k)fluoranthene	1.214	1.074	11.5	129	0.00
37 C	Benzo(a)pyrene	0.913	0.793	13.1	134	0.00
38	Dibenzo(a,h)anthracene	0.746	0.889	-19.2	189#	0.00
39	Benzo(g,h,i)perylene	0.830	0.898	-8.2	167#	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE072018\
Data File : BE097028.D
Acq On : 20 Jul 2018 15:36
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 20 17:02:54 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE071618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 20 11:34:20 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		