

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE072618\
 Data File : BE097111.D
 Acq On : 26 Jul 2018 20:20
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc : MDL 0.1PPM
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 27 00:50:47 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE072418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 24 13:24:10 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	77	0.00
2	1,4-Dioxane	0.400	0.459	-14.8	92	0.00
3	n-Nitrosodimethylamine	0.400	0.420	-5.0	87	0.00
4 S	2-Fluorophenol	0.400	0.358	10.5	74	0.00
5 S	Phenol-d6	0.400	0.365	8.8	76	0.00
6	bis(2-Chloroethyl)ether	0.400	0.434	-8.5	91	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	74	0.00
8 S	Nitrobenzene-d5	0.400	0.430	-7.5	85	0.00
9	Naphthalene	0.400	0.445	-11.2	87	0.00
10	Hexachlorobutadiene	0.400	0.466	-16.5	90	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.421	-5.2	82	0.00
12	2-Methylnaphthalene	0.400	0.425	-6.2	83	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	69	0.00
14 S	2,4,6-Tribromophenol	0.400	0.359	10.3	69	0.00
15 S	2-Fluorobiphenyl	0.400	0.469	-17.2	84	0.00
16	Acenaphthylene	0.400	0.447	-11.7	80	0.00
17	Acenaphthene	0.400	0.453	-13.2	81	0.00
18	Fluorene	0.400	0.452	-13.0	82	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	71	0.00
20	4-Bromophenyl-phenylether	0.400	0.432	-8.0	80	0.00
21	Hexachlorobenzene	0.400	0.481	-20.2	86	0.00
22	Pentachlorophenol	0.400	0.327	18.3	59	0.00
23	Phenanthrene	0.400	0.442	-10.5	82	0.00
24	Anthracene	0.400	0.405	-1.3	77	0.00
25 SURR	Fluoranthene-d10	0.400	0.399	0.3	77	0.00
26	Fluoranthene	0.400	0.418	-4.5	79	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	64	0.00
28	Pyrene	0.400	0.445	-11.2	75	0.00
29 S	Terphenyl-d14	0.400	0.451	-12.7	77	0.00
30	Benzo(a)anthracene	0.400	0.391	2.3	70	0.00
31	Chrysene	0.400	0.430	-7.5	71	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.329	17.8	54	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.376	6.0	64	0.00
34 I	Perylene-d12	0.400	0.400	0.0	66	0.00
35	Benzo(b)fluoranthene	0.400	0.355	11.3	63	0.00
36	Benzo(k)fluoranthene	0.400	0.443	-10.7	79	0.00
37 C	Benzo(a)pyrene	0.400	0.359	10.3	63	0.00
38	Dibenzo(a,h)anthracene	0.400	0.357	10.8	63	0.01
39	Benzo(g,h,i)perylene	0.400	0.341	14.8	60	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE072618\
Data File : BE097111.D
Acq On : 26 Jul 2018 20:20
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc : MDL 0.1PPM
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 27 00:50:47 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE072418.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 24 13:24:10 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				