

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE073118\
 Data File : BE097113.D
 Acq On : 31 Jul 2018 9:17
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 31 10:54:29 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE072418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 31 10:49:28 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	118	0.00
2	1,4-Dioxane	0.400	0.360	10.0	110	0.00
3	n-Nitrosodimethylamine	0.400	0.398	0.5	127	0.00
4 S	2-Fluorophenol	0.400	0.311	22.3	98	0.00
5 S	Phenol-d6	0.400	0.305	23.8	97	0.00
6	bis(2-Chloroethyl)ether	0.400	0.394	1.5	126	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	119	0.00
8 S	Nitrobenzene-d5	0.400	0.387	3.3	122	0.00
9	Naphthalene	0.400	0.397	0.8	124	0.00
10	Hexachlorobutadiene	0.400	0.416	-4.0	129	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.378	5.5	118	0.00
12	2-Methylnaphthalene	0.400	0.385	3.8	120	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	113	0.00
14 S	2,4,6-Tribromophenol	0.400	0.256	36.0#	80	0.00
15 S	2-Fluorobiphenyl	0.400	0.413	-3.2	120	0.00
16	Acenaphthylene	0.400	0.396	1.0	115	0.00
17	Acenaphthene	0.400	0.395	1.3	114	0.00
18	Fluorene	0.400	0.371	7.3	109	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	112	0.00
20	4-Bromophenyl-phenylether	0.400	0.398	0.5	117	0.00
21	Hexachlorobenzene	0.400	0.424	-6.0	122	0.00
22	Pentachlorophenol	0.400	0.379	5.3	112	0.00
23	Phenanthrene	0.400	0.396	1.0	116	0.00
24	Anthracene	0.400	0.370	7.5	112	0.00
25 SURR	Fluoranthene-d10	0.400	0.365	8.8	112	0.00
26	Fluoranthene	0.400	0.385	3.8	116	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	105	0.00
28	Pvrene	0.400	0.410	-2.5	113	0.00
29 S	Terphenyl-d14	0.400	0.410	-2.5	114	0.00
30	Benzo(a)anthracene	0.400	0.327	18.3	95	0.00
31	Chrysene	0.400	0.413	-3.2	111	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.245	38.8#	66	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.355	11.3	99	0.00
34 I	Perylene-d12	0.400	0.400	0.0	109	0.00
35	Benzo(b)fluoranthene	0.400	0.310	22.5	89	0.00
36	Benzo(k)fluoranthene	0.400	0.389	2.8	114	0.00
37 C	Benzo(a)pyrene	0.400	0.347	13.3	100	0.00
38	Dibenzo(a,h)anthracene	0.400	0.327	18.3	94	0.00
39	Benzo(a,h,i)perylene	0.400	0.340	15.0	97	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE073118\
Data File : BE097113.D
Acq On : 31 Jul 2018 9:17
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 31 10:54:29 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE072418.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 31 10:49:28 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				