

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE073118\
 Data File : BE097113.D
 Acq On : 31 Jul 2018 9:17
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 31 10:54:29 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE072418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 31 10:49:28 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	118	0.00
2	1,4-Dioxane	0.636	0.571	10.2	110	0.00
3	n-Nitrosodimethylamine	0.711	0.707	0.6	127	0.00
4 S	2-Fluorophenol	1.124	0.875	22.2	98	0.00
5 S	Phenol-d6	1.649	1.256	23.8	97	0.00
6	bis(2-Chloroethyl)ether	1.506	1.485	1.4	126	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	119	0.00
8 S	Nitrobenzene-d5	0.347	0.336	3.2	122	0.00
9	Naphthalene	3.590	3.567	0.6	124	0.00
10	Hexachlorobutadiene	0.161	0.167	-3.7	129	0.00
11 SURR	2-Methylnaphthalene-d10	0.599	0.566	5.5	118	0.00
12	2-Methylnaphthalene	0.616	0.593	3.7	120	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	113	0.00
14 S	2,4,6-Tribromophenol	0.132	0.084	36.4#	80	0.00
15 S	2-Fluorobiphenyl	1.381	1.427	-3.3	120	0.00
16	Acenaphthylene	1.795	1.778	0.9	115	0.00
17	Acenaphthene	1.044	1.030	1.3	114	0.00
18	Fluorene	1.379	1.278	7.3	109	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	112	0.00
20	4-Bromophenyl-phenylether	0.194	0.193	0.5	117	0.00
21	Hexachlorobenzene	0.209	0.221	-5.7	122	0.00
22	Pentachlorophenol	0.057	0.045	21.1	112	0.00
23	Phenanthrene	0.953	0.943	1.0	116	0.00
24	Anthracene	0.853	0.789	7.5	112	0.00
25 SURR	Fluoranthene-d10	4.003	3.648	8.9	112	0.00
26	Fluoranthene	1.070	1.030	3.7	116	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	105	0.00
28	Pvrene	1.097	1.125	-2.6	113	0.00
29 S	Terphenyl-d14	0.768	0.788	-2.6	114	0.00
30	Benzo(a)anthracene	0.995	0.813	18.3	95	0.00
31	Chrysene	1.067	1.101	-3.2	111	0.00
32	Bis(2-ethylhexyl)phthalate	1.243	0.638	48.7#	66	0.00
33	Indeno(1,2,3-cd)pyrene	0.966	0.857	11.3	99	0.00
34 I	Perylene-d12	1.000	1.000	0.0	109	0.00
35	Benzo(b)fluoranthene	1.015	0.787	22.5	89	0.00
36	Benzo(k)fluoranthene	1.185	1.152	2.8	114	0.00
37 C	Benzo(a)pyrene	0.955	0.828	13.3	100	0.00
38	Dibenzo(a,h)anthracene	0.953	0.779	18.3	94	0.00
39	Benzo(a,h,i)perylene	1.011	0.859	15.0	97	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE073118\
Data File : BE097113.D
Acq On : 31 Jul 2018 9:17
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 31 10:54:29 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE072418.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 31 10:49:28 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			