

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE073118\
 Data File : BE097123.D
 Acq On : 31 Jul 2018 15:46
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 31 16:51:25 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE072418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 31 10:49:28 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	144	0.00
2	1,4-Dioxane	0.636	0.580	8.8	135	0.00
3	n-Nitrosodimethylamine	0.711	0.698	1.8	152#	0.00
4 S	2-Fluorophenol	1.124	1.156	-2.8	157#	0.00
5 S	Phenol-d6	1.649	1.787	-8.4	168#	0.00
6	bis(2-Chloroethyl)ether	1.506	1.477	1.9	152#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	148	0.00
8 S	Nitrobenzene-d5	0.347	0.355	-2.3	160#	0.00
9	Naphthalene	3.590	3.488	2.8	151#	0.00
10	Hexachlorobutadiene	0.161	0.156	3.1	149	0.00
11 SURR	2-Methylnaphthalene-d10	0.599	0.584	2.5	151#	0.00
12	2-Methylnaphthalene	0.616	0.607	1.5	153#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	151#	0.00
14 S	2,4,6-Tribromophenol	0.132	0.148	-12.1	188#	0.00
15 S	2-Fluorobiphenyl	1.381	1.347	2.5	151#	0.00
16	Acenaphthylene	1.795	1.752	2.4	152#	0.00
17	Acenaphthene	1.044	1.032	1.1	153#	0.00
18	Fluorene	1.379	1.379	0.0	158#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	154#	0.00
20	4-Bromophenyl-phenylether	0.194	0.189	2.6	157#	0.00
21	Hexachlorobenzene	0.209	0.193	7.7	145	0.00
22	Pentachlorophenol	0.057	0.085	-49.1#	286#	0.01
23	Phenanthrene	0.953	0.933	2.1	158#	0.00
24	Anthracene	0.853	0.849	0.5	165#	0.01
25 SURR	Fluoranthene-d10	4.003	3.976	0.7	167#	0.00
26	Fluoranthene	1.070	1.059	1.0	163#	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	170#	0.00
28	Pvrene	1.097	1.064	3.0	173#	0.00
29 S	Terphenyl-d14	0.768	0.745	3.0	175#	0.00
30	Benzo(a)anthracene	0.995	0.994	0.1	189#	0.00
31	Chrysene	1.067	1.072	-0.5	176#	0.01
32	Bis(2-ethylhexyl)phthalate	1.243	1.833	-47.5#	306#	0.00
33	Indeno(1,2,3-cd)pyrene	0.966	1.150	-19.0	217#	0.00
34 I	Perylene-d12	1.000	1.000	0.0	204#	0.00
35	Benzo(b)fluoranthene	1.015	0.944	7.0	201#	0.00
36	Benzo(k)fluoranthene	1.185	1.079	8.9	200#	0.00
37 C	Benzo(a)pyrene	0.955	0.938	1.8	213#	0.00
38	Dibenzo(a,h)anthracene	0.953	0.958	-0.5	217#	0.00
39	Benzo(a,h,i)perylene	1.011	0.983	2.8	209#	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE073118\
Data File : BE097123.D
Acq On : 31 Jul 2018 15:46
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 31 16:51:25 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE072418.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 31 10:49:28 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			